

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Europastry **SDG No.:** Q2276
Contract: EURO03 **Lab Code:** CHEM **Case No.:** Q2276 **SAS No.:** Q2276
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	5.34			-20	20	06/11/2025	14:14	LB136119
	Copper	-6.25	2.0	312	-18	22	06/11/2025	14:14	LB136119
	Zinc	5.34			-40	40	06/11/2025	14:14	LB136119
ICSAB01	Arsenic	96.7	104	93	88.4	120	06/11/2025	14:27	LB136119
	Copper	444	511	87	434	588	06/11/2025	14:27	LB136119
	Zinc	992	952	104	809	1095	06/11/2025	14:27	LB136119
ICSA01	Arsenic	3.55			-20	20	06/12/2025	18:23	LB136140
	Copper	-10.4	2.0	521	-18	22	06/12/2025	18:23	LB136140
	Zinc	6.10			-40	40	06/12/2025	18:23	LB136140
ICSAB01	Arsenic	105	104	101	88.4	120	06/12/2025	18:34	LB136140
	Copper	446	511	87	434	588	06/12/2025	18:34	LB136140
	Zinc	1030	952	108	809	1095	06/12/2025	18:34	LB136140