

metals
- 5a -
MATRIX SPIKE SUMMARY

client: <u>Europastry</u>	level: <u>low</u>	sdg no.: <u>Q2276</u>
contract: <u>EURO03</u>	lab code: <u>CHEM</u>	case no.: <u>Q2276</u> sas no.: <u>Q2276</u>
matrix: <u>Water</u>	sample id: <u>Q2279-01</u>	client id: <u>301-469-5TH-AVEMS</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q2279-01MS</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Arsenic	ug/L	75 - 125	381		6.72	J	400	93		P
Copper	ug/L	75 - 125	254		117		150	91		P
Zinc	ug/L	75 - 125	270		166		100	104		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: Europastry **level:** low **sdg no.:** Q2276
contract: EURO03 **lab code:** CHEM **case no.:** Q2276 **sas no.:** Q2276
matrix: Water **sample id:** Q2279-01 **client id:** 301-469-5TH-AVEMSD
Percent Solids for Sample: NA **Spiked ID:** Q2279-01MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Arsenic	ug/L	75 - 125	374		6.72	J	400	92		P
Copper	ug/L	75 - 125	251		117		150	89		P
Zinc	ug/L	75 - 125	259		166		100	93		P