

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Europastry **Level:** LOW **SDG No.:** Q2276
Contract: EURO03 **Lab Code:** CHEM **Case No.:** Q2276 **SAS No.:** Q2276
Matrix: Water **Sample ID:** Q2279-01 **Client ID:** 301-469-5TH-AVEDUP
Percent Solids for Sample: NA **Duplicate ID** Q2279-01DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	6.72	J	8.39	J	22		P
Copper	ug/L	20	117		115		2		P
Zinc	ug/L	20	166		167		1		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Europastry **Level:** LOW **SDG No.:** Q2276
Contract: EURO03 **Lab Code:** CHEM **Case No.:** Q2276 **SAS No.:** Q2276
Matrix: Water **Sample ID:** Q2279-01MS **Client ID:** 301-469-5TH-AVEMSD
Percent Solids for Sample: NA **Duplicate ID** Q2279-01MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	381		374		2		P
Copper	ug/L	20	254		251		1		P
Zinc	ug/L	20	270		259		4		P