

DATA PACKAGE

GENERAL CHEMISTRY
METALS

PROJECT NAME : AEC-2025-0013- CSC 4019 WATERBURY

ATG-GREENVILLE AEC

426 Fairforest way

Greenville, SC - 29607

Phone No: 864-704-7984

ORDER ID : Q2289

ATTENTION : Daniel Maalouf



Laboratory Certification ID # 20012



1) Signature Page	3
2) Case Narrative	4
2.1) Metals-MS- Case Narrative	4
2.2) Genchem- Case Narrative	5
3) Qualifier Page	7
4) QA Checklist	8
5) Metals-MS Data	9
6) Genchem Data	15
7) Shipping Document	21
7.1) CHAIN OF CUSTODY	22
7.2) Lab Certificate	23

1
2
3
4
5
6
7

Cover Page

Order ID : Q2289

Project ID : AEC-2025-0013- CSC 4019 Waterbury

Client : ATG-GREENVILLE AEC

Lab Sample Number

Q2289-01
Q2289-02
Q2289-03

Client Sample Number

SP-1
SP-2
SP-3

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 6/19/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

ATG-GREENVILLE AEC

Project Name: AEC-2025-0013- CSC 4019 Waterbury

Project # N/A

Order ID # Q2289

Test Name: Metals ICP-Group1

A. Number of Samples and Date of Receipt:

3 Water samples were received on 06/11/2025.

B. Parameters:

According to the Chain of Custody document, the following analyses were requested: Anions Group1, COD, Metals ICP-Group1, Oil and Grease, pH, Phosphorus-Total, TKN and TSS. This data package contains results for Metals ICP-Group1.

C. Analytical Techniques:

The analysis of Metals ICP-Group1 was based on method 6020B and digestion based on method 3010 (waters).

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Blank Spike met requirements for all samples.

The Duplicate analysis met criteria for all samples.

The Matrix Spike analysis met criteria for all samples.

The Matrix Spike Duplicate analysis met criteria for all samples.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

The Serial Dilution met the acceptable requirements.

E. Additional Comments:

Collision cell is being used to remove potential interferences. The analytes Na, Mg, Al, K, V, Cr, Mn, Fe, Co, Ni, Cu, Zn, As are being analyzed with collision cell and analytes Be, B, Ca, Ti, Se, Sr, Zr, Mo, Ag, Cd, Sn, Sb, Ba, Tl, Pb, U are being analyzed with Non-Collision Cell. Helium gas is used for the Collision Cell analysis.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

CASE NARRATIVE

ATG-GREENVILLE AEC

Project Name: AEC-2025-0013- CSC 4019 Waterbury

Project # N/A

Order ID # Q2289

Test Name: Anions Group1,COD,Oil and Grease,pH,Phosphorus-Total,TKN,TSS

A. Number of Samples and Date of Receipt:

3 Water samples were received on 06/11/2025.

B. Parameters:

According to the Chain of Custody document, the following analyses were requested: Anions Group1, COD, Metals ICP-Group1, Oil and Grease, pH, Phosphorus-Total, TKN and TSS. This data package contains results for Anions Group1,COD, Oil and Grease, pH, Phosphorus-Total, TKN, TSS.

C. Analytical Techniques:

The analysis of Oil and Grease was based on method 1664A, The analysis of Anions Group1 was based on method 300.0, The analysis of Phosphorus-Total was based on method 365.3, The analysis of pH was based on method 9040C, The analysis of TSS was based on method SM2540 D, The analysis of TKN was based on method SM4500 N Org B or C and The analysis of COD was based on method SM5220 D.

D. QA/ QC Samples:

The Holding Times were met for all samples except for SP-1 of pH, for SP-2 of pH. for SP-3 of pH as samples were receive out of holding time.

The Blank Spike met requirements for all samples.

The Duplicate analysis met criteria for all samples.

The Matrix Spike (LAW-25-0084MS) analysis met criteria for all elements except for TKN due to matrix interference.

The Matrix Spike Duplicate (LAW-25-0084MSD) analysis met criteria for all elements except for TKN due to matrix interference.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

E. Additional Comments:

As per method 1664A, MS/MSD is required to be performed with the sample analysis. However, Lab did not receive sufficient volume to perform the MS/MSD therefore MS/MSD was not performed for this project.

Sample Q2289-01 and Q2289-03 analyzed straight dilution for COD because, the original samples were reading over range, only 5X and 10 X dilutions has been reported.



I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

DATA REPORTING QUALIFIERS- INORGANIC

For reporting results, the following “ Results Qualifiers” are used:

J	Indicates the reported value was obtained from a reading that was less than the Contract Required Detection Limit (CRDL), but greater than or equal to the Instrument Detection Limit (IDL).
U	Indicates the analyte was analyzed for, but not detected.
ND	Indicates the analyte was analyzed for, but not detected
E	Indicates the reported value is estimated because of the presence of interference
M	Indicates Duplicate injection precision not met.
N	Indicates the spiked sample recovery is not within control limits.
S	Indicates the reported value was determined by the Method of Standard Addition (MSA).
*	Indicates that the duplicate analysis is not within control limits.
+	Indicates the correlation coefficient for the MSA is less than 0.995.
D	Indicates the reported value is from a secondary analysis with a dilution factor. The original analysis exceeded the calibration range.
M	Method qualifiers “P” for ICP instrument “PM” for ICP when Microwave Digestion is used “CV” for Manual Cold Vapor AA “AV” for automated Cold Vapor AA “CA” for MIDI-Distillation Spectrophotometric “AS” for Semi -Automated Spectrophotometric “C” for Manual Spectrophotometric “T” for Titrimetric “NR” for analyte not required to be analyzed
OR	Indicates the analyte’s concentration exceeds the calibrated range of the instrument for that specific analysis.
Q	Indicates the LCS did not meet the control limits requirements
H	Sample Analysis Out Of Hold Time

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q2289

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: SOHIL JODHANI

Date: 06/19/2025

Hit Summary Sheet
SW-846

SDG No.:	Q2289	Order ID:	Q2289
Client:	ATG-GREENVILLE AEC	Project ID:	AEC-2025-0013- CSC 4019 Waterbury

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID : SP-1								
Q2289-01	SP-1	Water	Copper	103		0.30	2.00	ug/L
Q2289-01	SP-1	Water	Lead	41.2		0.21	1.00	ug/L
Q2289-01	SP-1	Water	Zinc	376		1.25	5.00	ug/L
Client ID : SP-2								
Q2289-02	SP-2	Water	Copper	59.5		0.30	2.00	ug/L
Q2289-02	SP-2	Water	Lead	25.5		0.21	1.00	ug/L
Q2289-02	SP-2	Water	Zinc	167		1.25	5.00	ug/L
Client ID : SP-3								
Q2289-03	SP-3	Water	Copper	335		0.30	2.00	ug/L
Q2289-03	SP-3	Water	Lead	87.8		0.21	1.00	ug/L
Q2289-03	SP-3	Water	Zinc	830		1.25	5.00	ug/L



SAMPLE DATA

Report of Analysis

Client:	ATG-GREENVILLE AEC	Date Collected:	06/10/25
Project:	AEC-2025-0013- CSC 4019 Waterbury	Date Received:	06/11/25
Client Sample ID:	SP-1	SDG No.:	Q2289
Lab Sample ID:	Q2289-01	Matrix:	Water
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7440-50-8	Copper	103		1	0.30	2.00	ug/L	06/17/25 10:10	06/18/25 12:59	6020B	3010A
7439-92-1	Lead	41.2		1	0.21	1.00	ug/L	06/17/25 10:10	06/18/25 12:59	6020B	3010A
7440-66-6	Zinc	376		1	1.25	5.00	ug/L	06/17/25 10:10	06/18/25 12:59	6020B	3010A

Color Before:	light Brown	Clarity Before:	Cloudy	Texture:
Color After:	light Brown	Clarity After:	Clear	Artifacts:
Comments:	Metals ICP-Group1			

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N =Spiked sample recovery not within control limits

Report of Analysis

Client:	ATG-GREENVILLE AEC	Date Collected:	06/10/25
Project:	AEC-2025-0013- CSC 4019 Waterbury	Date Received:	06/11/25
Client Sample ID:	SP-2	SDG No.:	Q2289
Lab Sample ID:	Q2289-02	Matrix:	Water
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7440-50-8	Copper	59.5	1	0.30		2.00	ug/L	06/17/25 10:10	06/18/25 13:02	6020B	3010A
7439-92-1	Lead	25.5	1	0.21		1.00	ug/L	06/17/25 10:10	06/18/25 13:02	6020B	3010A
7440-66-6	Zinc	167	1	1.25		5.00	ug/L	06/17/25 10:10	06/18/25 13:02	6020B	3010A

Color Before:	light Brown	Clarity Before:	Cloudy	Texture:
Color After:	light Brown	Clarity After:	Clear	Artifacts:
Comments:	Metals ICP-Group1			

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N =Spiked sample recovery not within control limits

Report of Analysis

Client:	ATG-GREENVILLE AEC	Date Collected:	06/10/25
Project:	AEC-2025-0013- CSC 4019 Waterbury	Date Received:	06/11/25
Client Sample ID:	SP-3	SDG No.:	Q2289
Lab Sample ID:	Q2289-03	Matrix:	Water
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7440-50-8	Copper	335		1	0.30	2.00	ug/L	06/17/25 10:10	06/18/25 13:05	6020B	3010A
7439-92-1	Lead	87.8		1	0.21	1.00	ug/L	06/17/25 10:10	06/18/25 13:05	6020B	3010A
7440-66-6	Zinc	830		1	1.25	5.00	ug/L	06/17/25 10:10	06/18/25 13:05	6020B	3010A

Color Before:	Brown	Clarity Before:	Cloudy	Texture:
Color After:	light Brown	Clarity After:	Clear	Artifacts:
Comments:	Metals ICP-Group1			

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N =Spiked sample recovery not within control limits

LAB CHRONICLE

OrderID:	Q2289	OrderDate:	6/11/2025 11:44:58 AM
Client:	ATG-GREENVILLE AEC	Project:	AEC-2025-0013- CSC 4019 Waterbury
Contact:	Daniel Maalouf	Location:	A11

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2289-01	SP-1	Water	Metals ICP-Group1	6020B	06/10/25	06/17/25	06/18/25	06/11/25
Q2289-02	SP-2	Water	Metals ICP-Group1	6020B	06/10/25	06/17/25	06/18/25	06/11/25
Q2289-03	SP-3	Water	Metals ICP-Group1	6020B	06/10/25	06/17/25	06/18/25	06/11/25



SAMPLE DATA

Report of Analysis

Client:	ATG-GREENVILLE AEC	Date Collected:	06/10/25 10:50
Project:	AEC-2025-0013- CSC 4019 Waterbury	Date Received:	06/11/25
Client Sample ID:	SP-1	SDG No.:	Q2289
Lab Sample ID:	Q2289-01	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Nitrite	0.084	J	1	0.074	0.60	mg/L		06/11/25 08:43	300.0
Nitrate	0.55		1	0.095	0.50	mg/L		06/11/25 08:43	300.0
Nitrate+Nitrite	0.63	J	1	0.17	1.10	mg/L		06/11/25 08:43	300.0
COD	229	D	5	7.50	50.0	mg/L		06/12/25 14:18	SM 5220 D-11
Oil and Grease	5.70		1	0.29	5.00	mg/L		06/17/25 09:30	1664A
pH	7.59	H	1	0	0	pH		06/12/25 08:40	9040C
Phosphorus, Total	0.046	J	1	0.0050	0.050	mg/L	06/13/25 09:45	06/13/25 13:22	365.3
TKN	2.50		1	0.11	0.50	mg/L	06/12/25 08:50	06/16/25 15:42	SM4500-N Org C-11 plus NH3 B plus G-11
TSS	600		1	1.00	4.00	mg/L		06/13/25 10:00	SM 2540 D-15

Comments: pH result reported at temperature 20.6 °C

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

Report of Analysis

Client:	ATG-GREENVILLE AEC	Date Collected:	06/10/25 10:26
Project:	AEC-2025-0013- CSC 4019 Waterbury	Date Received:	06/11/25
Client Sample ID:	SP-2	SDG No.:	Q2289
Lab Sample ID:	Q2289-02	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Nitrite	0.074	U	1	0.074	0.60	mg/L		06/11/25 09:04	300.0
Nitrate	0.23	J	1	0.095	0.50	mg/L		06/11/25 09:04	300.0
Nitrate+Nitrite	0.23	J	1	0.17	1.10	mg/L		06/11/25 09:04	300.0
COD	131		1	1.50	10.0	mg/L		06/12/25 14:19	SM 5220 D-11
Oil and Grease	7.80		1	0.29	5.00	mg/L		06/17/25 09:30	1664A
pH	7.34	H	1	0	0	pH		06/12/25 08:46	9040C
Phosphorus, Total	0.062		1	0.0050	0.050	mg/L	06/13/25 09:45	06/13/25 13:24	365.3
TKN	1.50		1	0.11	0.50	mg/L	06/12/25 08:50	06/16/25 15:43	SM4500-N Org C-11 plus NH3 B plus G-11
TSS	305		1	1.00	4.00	mg/L		06/13/25 10:00	SM 2540 D-15

Comments: pH result reported at temperature 20.7 °C

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

Report of Analysis

Client:	ATG-GREENVILLE AEC	Date Collected:	06/10/25 10:33
Project:	AEC-2025-0013- CSC 4019 Waterbury	Date Received:	06/11/25
Client Sample ID:	SP-3	SDG No.:	Q2289
Lab Sample ID:	Q2289-03	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Nitrite	0.085	J	1	0.074	0.60	mg/L		06/11/25 09:26	300.0
Nitrate	0.27	J	1	0.095	0.50	mg/L		06/11/25 09:26	300.0
Nitrate+Nitrite	0.36	J	1	0.17	1.10	mg/L		06/11/25 09:26	300.0
COD	774	D	10	15.0	100	mg/L		06/12/25 14:19	SM 5220 D-11
Oil and Grease	11.3		1	0.29	5.00	mg/L		06/17/25 09:30	1664A
pH	7.78	H	1	0	0	pH		06/12/25 08:50	9040C
Phosphorus, Total	0.094		1	0.0050	0.050	mg/L	06/13/25 09:45	06/13/25 13:25	365.3
TKN	5.50		1	0.11	0.50	mg/L	06/12/25 08:50	06/16/25 15:43	SM4500-N Org C-11 plus NH3 B plus G-11
TSS	1040		1	1.00	4.00	mg/L		06/13/25 10:00	SM 2540 D-15

Comments: pH result reported at temperature 20.8 °C

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

LAB CHRONICLE

OrderID:	Q2289	OrderDate:	6/11/2025 11:44:58 AM
Client:	ATG-GREENVILLE AEC	Project:	AEC-2025-0013- CSC 4019 Waterbury
Contact:	Daniel Maalouf	Location:	A11

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2289-01	SP-1	WATER			06/10/25 10:50			06/11/25
			Anions Group1	300.0			06/11/25 08:43	
			COD	SM5220 D			06/12/25 14:18	
			Oil and Grease	1664A			06/17/25 09:30	
			pH	9040C			06/12/25 08:40	
			Phosphorus-Total	365.3		06/13/25	06/13/25 13:22	
			TKN	SM4500-N Org C-11 plus NH3 B plus G-11		06/12/25	06/16/25 15:42	
			TSS	SM2540 D			06/13/25 10:00	
Q2289-02	SP-2	WATER			06/10/25 10:26			06/11/25
			Anions Group1	300.0			06/11/25 09:04	
			COD	SM5220 D			06/12/25 14:19	
			Oil and Grease	1664A			06/17/25 09:30	
			pH	9040C			06/12/25 08:46	
			Phosphorus-Total	365.3		06/13/25	06/13/25 13:24	


LAB CHRONICLE

Q2289-03	SP-3	WATER	TKN	SM4500-N Org C-11 plus NH3 B plus G-11	06/12/25	06/16/25 15:43	06/10/25 10:33	06/11/25
			TSS	SM2540 D		06/13/25 10:00		
			Anions Group1	300.0		06/11/25 09:26		
			COD	SM5220 D		06/12/25 14:19		
			Oil and Grease	1664A		06/17/25 09:30		
			pH	9040C		06/12/25 08:50		
			Phosphorus-Total	365.3	06/13/25	06/13/25 13:25		
			TKN	SM4500-N Org C-11 plus NH3 B plus G-11	06/12/25	06/16/25 15:43		
			TSS	SM2540 D		06/13/25 10:00		



SHIPPING DOCUMENTS

Q 2289



284 Sheffield Street, Mountainside, NJ 07092
(908) 789-8900 Fax: (908) 788-9222
www.chemtech.net

CHAIN OF CUSTODY RECORD

Alliance Project Number: AEC-2025-0013

COC Number:

CLIENT INFORMATION			PROJECT INFORMATION				BILLING INFORMATION										
COMPANY: Durham School Services			PROJECT NAME: CSC 4019 Waterbury				BILL TO: candice.caviness@alliancetg.com PO#										
ADDRESS: 62 Fulkerson Dr			PROJECT #: AEC-2025-0013 LOCATION: Waterbury				ADDRESS:										
CITY: Waterbury STATE: CT ZIP: 29615			PROJECT MANAGER:				CITY: STATE: ZIP:										
ATTENTION: daniel.maalouf@alliancetg.com			E-MAIL:				ATTENTION: PHONE:										
PHONE: FAX:			PHONE: 864-704-7984 FAX:				ANALYSIS										
DATA TURNAROUND INFORMATION			DATA DELIVERABLE INFORMATION				PRESERVATIVES										COMMENTS
FAX: DAYS* HARD COPY: DAYS* EDD: DAYS* * TO BE APPROVED BY ALLIANCE STANDARD TURNAROUND TIME IS 10 BUSINESS DAYS			<input type="checkbox"/> RESULTS ONLY <input type="checkbox"/> USEPA CLP <input type="checkbox"/> RESULTS + QC <input type="checkbox"/> New York State ASP "B" <input type="checkbox"/> New Jersey REDUCED <input type="checkbox"/> New York State ASP "A" <input type="checkbox"/> New Jersey CLP <input type="checkbox"/> Other <input type="checkbox"/> EDD Format														
CHEMTECH SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# of Bottles										
			COMP	GRAB	DATE	TIME		1	2	3	4	5	6	7	8	9	
1.	SP-1	SW		X		10:50	5	1	1	1	1	1	1	1	1		
2.	SP-2	SW		X		10:26	5	1	1	1	1	1	1	1	1		
3.	SP-3	SW		X		10:33	5	1	1	1	1	1	1	1	1		
4.																	
5.																	
6.																	
7.																	
8.																	
9.																	
10.																	

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE PROSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER	DATE/TIME	RECEIVED BY	Conditions of bottles or coolers at receipt: <input type="checkbox"/> Compliant <input type="checkbox"/> Non Compliant <input type="checkbox"/> Cooler Temp 3.0 MeOH extraction requires an additional 4oz. Jar for percent solid Comments:		
1. Daniel Maalouf ATG		1.			
RELINQUISHED BY	DATE/TIME 6/11/25 12:35	RECEIVED BY 2. [Signature]			
RELINQUISHED BY	DATE/TIME	RECEIVED FOR LAB BY	Page of	SHIPPED VIA: CLIENT: <input type="checkbox"/> Hand Delivered <input type="checkbox"/> Overnight ALLIANCE: <input type="checkbox"/> Picked Up <input type="checkbox"/> Overnight	Shipment Complete <input type="checkbox"/> YES <input type="checkbox"/> NO
3.		3. [Signature]			

WHITE - ALLIANCE COPY FOR RETURN TO CLIENT YELLOW - ALLIANCE COPY PINK - SAMPLER COPY

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488