

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ATG-GREENVILLE AEC **SDG No.:** Q2289
Contract: ATGG01 **Lab Code:** CHEM **Case No.:** Q2289 **SAS No.:** Q2289
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV1	Copper	101	100	100	90 - 110	P	06/18/2025	12:01	LB136195
	Lead	213	200	106	90 - 110	P	06/18/2025	12:01	LB136195
	Zinc	202	200	101	90 - 110	P	06/18/2025	12:01	LB136195

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ATG-GREENVILLE AEC **SDG No.:** Q2289
Contract: ATGG01 **Lab Code:** CHEM **Case No.:** Q2289 **SAS No.:** Q2289
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV1	Copper	2.09	2.0	104	80 - 120	P	06/18/2025	12:08	LB136195
	Lead	0.92	1.0	92	80 - 120	P	06/18/2025	12:08	LB136195
	Zinc	5.36	5.0	107	80 - 120	P	06/18/2025	12:08	LB136195

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ATG-GREENVILLE AEC SDG No.: Q2289
 Contract: ATGG01 Lab Code: CHEM Case No.: Q2289 SAS No.: Q2289
 Initial Calibration Source: EPA
 Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV1	Copper	5090	5000	102	90 - 110	P	06/18/2025	12:48	LB136195
	Lead	2670	2500	107	90 - 110	P	06/18/2025	12:48	LB136195
	Zinc	4980	5000	100	90 - 110	P	06/18/2025	12:48	LB136195
CCV2	Copper	5110	5000	102	90 - 110	P	06/18/2025	13:12	LB136195
	Lead	2640	2500	106	90 - 110	P	06/18/2025	13:12	LB136195
	Zinc	5010	5000	100	90 - 110	P	06/18/2025	13:12	LB136195
CCV3	Copper	5100	5000	102	90 - 110	P	06/18/2025	13:49	LB136195
	Lead	2630	2500	105	90 - 110	P	06/18/2025	13:49	LB136195
	Zinc	4960	5000	99	90 - 110	P	06/18/2025	13:49	LB136195
CCV4	Copper	5160	5000	103	90 - 110	P	06/18/2025	14:15	LB136195
	Lead	2600	2500	104	90 - 110	P	06/18/2025	14:15	LB136195
	Zinc	4980	5000	100	90 - 110	P	06/18/2025	14:15	LB136195