



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: ATG-GREENVILLE AEC **SDG No.:** Q2289
Contract: ATGG01 **Lab Code:** CHEM **Case No.:** Q2289 **SAS No.:** Q2289
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI	Copper	2.43	2.0	122	70 - 130	P	06/18/2025	12:55	LB136195
	Lead	0.95	1.0	95	70 - 130	P	06/18/2025	12:55	LB136195
	Zinc	5.03	5.0	101	50 - 150	P	06/18/2025	12:55	LB136195