



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ATG-GREENVILLE AEC **SDG No.:** Q2289
Contract: ATGG01 **Lab Code:** CHEM **Case No.:** Q2289 **SAS No.:** Q2289

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB1	Copper	0.30	+/-1	U	2.00	P	06/18/2025	12:13	LB136195
	Lead	0.21	+/-0.5	U	1.00	P	06/18/2025	12:13	LB136195
	Zinc	1.25	+/-2.5	U	5.00	P	06/18/2025	12:13	LB136195

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ATG-GREENVILLE AEC **SDG No.:** Q2289
Contract: ATGG01 **Lab Code:** CHEM **Case No.:** Q2289 **SAS No.:** Q2289

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB1	Copper	0.48	+/-1	J	2.00	P	06/18/2025	12:51	LB136195
	Lead	0.21	+/-0.5	U	1.00	P	06/18/2025	12:51	LB136195
	Zinc	1.25	+/-2.5	U	5.00	P	06/18/2025	12:51	LB136195
CCB2	Copper	0.30	+/-1	U	2.00	P	06/18/2025	13:15	LB136195
	Lead	0.21	+/-0.5	U	1.00	P	06/18/2025	13:15	LB136195
	Zinc	1.25	+/-2.5	U	5.00	P	06/18/2025	13:15	LB136195
CCB3	Copper	0.30	+/-1	U	2.00	P	06/18/2025	13:52	LB136195
	Lead	0.21	+/-0.5	U	1.00	P	06/18/2025	13:52	LB136195
	Zinc	1.25	+/-2.5	U	5.00	P	06/18/2025	13:52	LB136195
CCB4	Copper	0.30	+/-1	U	2.00	P	06/18/2025	14:18	LB136195
	Lead	0.21	+/-0.5	U	1.00	P	06/18/2025	14:18	LB136195
	Zinc	1.25	+/-2.5	U	5.00	P	06/18/2025	14:18	LB136195