

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: ATG-GREENVILLE AEC **SDG No.:** Q2289
Contract: ATGG01 **Lab Code:** CHEM **Case No.:** Q2289 **SAS No.:** Q2289
ICS Source: EPA **Instrument ID:** P7

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA1	Copper	9.31	8.0	116	4	12	06/18/2025	12:30	LB136195
	Lead	4.06	4.0	102	2	6	06/18/2025	12:30	LB136195
	Zinc	11.6	11.0	106	1	21	06/18/2025	12:30	LB136195
ICSAB1	Copper	27.1	25.0	108	21	29	06/18/2025	12:45	LB136195
	Lead	24.4	25.0	98	21.3	28.8	06/18/2025	12:45	LB136195
	Zinc	30.4	29.0	105	19	39	06/18/2025	12:45	LB136195