

metals
- 5a -
MATRIX SPIKE SUMMARY

client: <u>ATG-GREENVILLE AEC</u>	level: <u>low</u>	sdg no.: <u>Q2289</u>
contract: <u>ATGG01</u>	lab code: <u>CHEM</u>	case no.: <u>Q2289</u> sas no.: <u>Q2289</u>
matrix: <u>Water</u>	sample id: <u>Q2289-02</u>	client id: <u>SP-2MS</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q2289-02MS</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	75 - 125	5170		59.5		5000	102		P
Lead	ug/L	75 - 125	2510		25.5		2500	99		P
Zinc	ug/L	75 - 125	5440		167		5000	105		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: <u>ATG-GREENVILLE AEC</u>	level: <u>low</u>	sdg no.: <u>Q2289</u>
contract: <u>ATGG01</u>	lab code: <u>CHEM</u>	case no.: <u>Q2289</u> sas no.: <u>Q2289</u>
matrix: <u>Water</u>	sample id: <u>Q2289-02</u>	client id: <u>SP-2MSD</u>
Percent Solids for Sample: NA	Spiked ID: Q2289-02MSD	Percent Solids for Spike Sample: NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	75 - 125	5080		59.5		5000	100		P
Lead	ug/L	75 - 125	2470		25.5		2500	98		P
Zinc	ug/L	75 - 125	5380		167		5000	104		P