



Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: ATG-GREENVILLE AEC **SDG No.:** Q2289
Contract: ATGG01 **Lab Code:** CHEM **Case No.:** Q2289 **SAS No.:** Q2289

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
<hr/>							
PB168510BS							
Copper	ug/L	5000	5190		104	80 - 120	P
Lead	ug/L	2500	2480		99	80 - 120	P
Zinc	ug/L	5000	5280		106	80 - 120	P
