

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ATG-GREENVILLE AEC **SDG No.:** Q2290
Contract: ATGG01 **Lab Code:** CHEM **Case No.:** Q2290 **SAS No.:** Q2290
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Copper	1050	1000	104	95 - 105	P	06/17/2025	12:25	LB136187
	Lead	3890	4000	97	95 - 105	P	06/17/2025	12:25	LB136187
	Zinc	2020	2000	101	95 - 105	P	06/17/2025	12:25	LB136187

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ATG-GREENVILLE AEC **SDG No.:** Q2290
Contract: ATGG01 **Lab Code:** CHEM **Case No.:** Q2290 **SAS No.:** Q2290
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Copper	23.3	20.0	117	80 - 120	P	06/17/2025	12:41	LB136187
	Lead	12.5	12.0	104	80 - 120	P	06/17/2025	12:41	LB136187
	Zinc	41.5	40.0	104	80 - 120	P	06/17/2025	12:41	LB136187

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ATG-GREENVILLE AEC **SDG No.:** Q2290

Contract: ATGG01 **Lab Code:** CHEM **Case No.:** Q2290 **SAS No.:** Q2290

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Copper	1250	1250	100	90 - 110	P	06/17/2025	13:40	LB136187
	Lead	4790	5000	96	90 - 110	P	06/17/2025	13:40	LB136187
	Zinc	2430	2500	97	90 - 110	P	06/17/2025	13:40	LB136187
CCV02	Copper	1260	1250	101	90 - 110	P	06/17/2025	14:43	LB136187
	Lead	4900	5000	98	90 - 110	P	06/17/2025	14:43	LB136187
	Zinc	2510	2500	100	90 - 110	P	06/17/2025	14:43	LB136187
CCV03	Copper	1230	1250	99	90 - 110	P	06/17/2025	15:38	LB136187
	Lead	4740	5000	95	90 - 110	P	06/17/2025	15:38	LB136187
	Zinc	2420	2500	97	90 - 110	P	06/17/2025	15:38	LB136187
CCV04	Copper	1290	1250	103	90 - 110	P	06/17/2025	16:30	LB136187
	Lead	4850	5000	97	90 - 110	P	06/17/2025	16:30	LB136187
	Zinc	2490	2500	100	90 - 110	P	06/17/2025	16:30	LB136187
CCV05	Copper	1280	1250	102	90 - 110	P	06/17/2025	17:37	LB136187
	Lead	4850	5000	97	90 - 110	P	06/17/2025	17:37	LB136187
	Zinc	2480	2500	99	90 - 110	P	06/17/2025	17:37	LB136187
CCV06	Copper	1270	1250	102	90 - 110	P	06/17/2025	18:38	LB136187
	Lead	4820	5000	96	90 - 110	P	06/17/2025	18:38	LB136187
	Zinc	2420	2500	97	90 - 110	P	06/17/2025	18:38	LB136187
CCV07	Copper	1300	1250	104	90 - 110	P	06/17/2025	19:25	LB136187
	Lead	4870	5000	97	90 - 110	P	06/17/2025	19:25	LB136187
	Zinc	2470	2500	99	90 - 110	P	06/17/2025	19:25	LB136187
CCV08	Copper	1270	1250	101	90 - 110	P	06/17/2025	20:17	LB136187
	Lead	4870	5000	98	90 - 110	P	06/17/2025	20:17	LB136187
	Zinc	2440	2500	98	90 - 110	P	06/17/2025	20:17	LB136187