



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: ATG-GREENVILLE AEC **SDG No.:** Q2290
Contract: ATGG01 **Lab Code:** CHEM **Case No.:** Q2290 **SAS No.:** Q2290
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Copper	23.2	20.0	116	65 - 135	P	06/17/2025	12:50	LB136187
	Lead	11.7	12.0	97	65 - 135	P	06/17/2025	12:50	LB136187
	Zinc	41.8	40.0	104	65 - 135	P	06/17/2025	12:50	LB136187