



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

### Metals

- 3a -

#### INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ATG-GREENVILLE AEC

SDG No.: Q2290

Contract: ATGG01

Lab Code: CHEM

Case No.: Q2290

SAS No.: Q2290

| Sample ID | Analyte | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|---------|----------------|---------------------|--------------|------|---|------------------|------------------|---------------|
| ICB01     | Copper  | 4.60           | +/-10               | U            | 20.0 | P | 06/17/2025       | 12:46            | LB136187      |
|           | Lead    | 2.30           | +/-6                | U            | 12.0 | P | 06/17/2025       | 12:46            | LB136187      |
|           | Zinc    | 3.50           | +/-20               | U            | 40.0 | P | 06/17/2025       | 12:46            | LB136187      |

# Metals

- 3a -

## INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ATG-GREENVILLE AEC SDG No.: Q2290  
Contract: ATGG01 Lab Code: CHEM Case No.: Q2290 SAS No.: Q2290

| Sample ID | Analyte | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|---------|----------------|---------------------|--------------|------|---|------------------|------------------|---------------|
| CCB01     | Copper  | 4.60           | +/-10               | U            | 20.0 | P | 06/17/2025       | 13:48            | LB136187      |
|           | Lead    | 2.30           | +/-6                | U            | 12.0 | P | 06/17/2025       | 13:48            | LB136187      |
|           | Zinc    | 3.50           | +/-20               | U            | 40.0 | P | 06/17/2025       | 13:48            | LB136187      |
| CCB02     | Copper  | 4.60           | +/-10               | U            | 20.0 | P | 06/17/2025       | 14:50            | LB136187      |
|           | Lead    | 2.30           | +/-6                | U            | 12.0 | P | 06/17/2025       | 14:50            | LB136187      |
|           | Zinc    | 3.50           | +/-20               | U            | 40.0 | P | 06/17/2025       | 14:50            | LB136187      |
| CCB03     | Copper  | 4.60           | +/-10               | U            | 20.0 | P | 06/17/2025       | 15:42            | LB136187      |
|           | Lead    | 2.30           | +/-6                | U            | 12.0 | P | 06/17/2025       | 15:42            | LB136187      |
|           | Zinc    | 3.50           | +/-20               | U            | 40.0 | P | 06/17/2025       | 15:42            | LB136187      |
| CCB04     | Copper  | 4.60           | +/-10               | U            | 20.0 | P | 06/17/2025       | 16:34            | LB136187      |
|           | Lead    | 2.30           | +/-6                | U            | 12.0 | P | 06/17/2025       | 16:34            | LB136187      |
|           | Zinc    | 3.50           | +/-20               | U            | 40.0 | P | 06/17/2025       | 16:34            | LB136187      |
| CCB05     | Copper  | 4.60           | +/-10               | U            | 20.0 | P | 06/17/2025       | 17:41            | LB136187      |
|           | Lead    | 2.94           | +/-6                | J            | 12.0 | P | 06/17/2025       | 17:41            | LB136187      |
|           | Zinc    | 3.50           | +/-20               | U            | 40.0 | P | 06/17/2025       | 17:41            | LB136187      |
| CCB06     | Copper  | 4.60           | +/-10               | U            | 20.0 | P | 06/17/2025       | 18:42            | LB136187      |
|           | Lead    | 2.30           | +/-6                | U            | 12.0 | P | 06/17/2025       | 18:42            | LB136187      |
|           | Zinc    | 3.50           | +/-20               | U            | 40.0 | P | 06/17/2025       | 18:42            | LB136187      |
| CCB07     | Copper  | 4.60           | +/-10               | U            | 20.0 | P | 06/17/2025       | 19:30            | LB136187      |
|           | Lead    | 2.30           | +/-6                | U            | 12.0 | P | 06/17/2025       | 19:30            | LB136187      |
|           | Zinc    | 3.50           | +/-20               | U            | 40.0 | P | 06/17/2025       | 19:30            | LB136187      |
| CCB08     | Copper  | 4.60           | +/-10               | U            | 20.0 | P | 06/17/2025       | 20:22            | LB136187      |
|           | Lead    | 4.53           | +/-6                | J            | 12.0 | P | 06/17/2025       | 20:22            | LB136187      |
|           | Zinc    | 4.70           | +/-20               | J            | 40.0 | P | 06/17/2025       | 20:22            | LB136187      |