



- 7 -

Client:	<u>ATG-GREENVILLE AEC</u>	SDG No.:	<u>Q2290</u>				
Contract:	ATGG01	Lab Code:	CHEM	Case No.:	Q2290	SAS No.:	Q2290

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168444BS							
Copper	ug/L	150	168		112	85 - 115	P
Lead	ug/L	500	471		94	85 - 115	P
Zinc	ug/L	100	104		104	85 - 115	P