

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: ATG-GREENVILLE AEC

SDG No.: Q2291

Contract: ATGG01

Lab Code: CHEM

Case No.: Q2291

SAS No.: Q2291

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168444BS							
Copper	ug/L	150	168		112	85 - 115	P
Lead	ug/L	500	471		94	85 - 115	P
Zinc	ug/L	100	104		104	85 - 115	P