

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB136131

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER, COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.05].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	06/12/2025	08:15
2	CAL2	1	Water	NA	NA	20.2	7.00	06/12/2025	08:16
3	CAL3	1	Water	NA	NA	20.1	10.02	06/12/2025	08:20
4	ICV	1	Water	NA	NA	20.2	7.02	06/12/2025	08:25
5	CCV1	1	Water	NA	NA	20.2	2.01	06/12/2025	08:33
6	Q2289-01	1	Water	NA	NA	20.6	7.59	06/12/2025	08:40
7	Q2289-01DUP	1	Water	NA	NA	20.8	7.61	06/12/2025	08:42
8	Q2289-02	1	Water	NA	NA	20.7	7.34	06/12/2025	08:46
9	Q2289-03	1	Water	NA	NA	20.8	7.78	06/12/2025	08:50
10	Q2290-01	1	Water	NA	NA	20.5	6.92	06/12/2025	09:00
11	Q2290-02	1	Water	NA	NA	20.1	7.09	06/12/2025	09:10
12	Q2291-01	1	Water	NA	NA	20.3	7.21	06/12/2025	09:15
13	Q2292-01	1	Water	NA	NA	20.1	7.17	06/12/2025	09:25
14	Q2292-02	1	Water	NA	NA	20.8	8.69	06/12/2025	09:30
15	Q2293-01	1	Water	NA	NA	20.3	7.31	06/12/2025	09:40
16	CCV2	1	Water	NA	NA	20.2	12.02	06/12/2025	09:47

WORKLIST(Hardcopy Internal Chain)

13136131

WorkList Name : PH S Q2291 WorkList ID : 190120 Department : Wet-Chemistry Date : 06-12-2025 07:59:33

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2289-01	SP-1	Water	pH	Cool 4 deg C	ATGG01	A11	06/10/2025	9040C
Q2289-02	SP-2	Water	pH	Cool 4 deg C	ATGG01	A11	06/10/2025	9040C
Q2289-03	SP-3	Water	pH	Cool 4 deg C	ATGG01	A11	06/10/2025	9040C
Q2290-01	Outfall 1	Water	pH	Cool 4 deg C	ATGG01	A11	06/10/2025	9040C
Q2290-02	Outfall 2	Water	pH	Cool 4 deg C	ATGG01	A11	06/10/2025	9040C
Q2291-01	Outfall 001	Water	pH	Cool 4 deg C	ATGG01	A11	06/10/2025	9040C
Q2292-01	SW-1	Water	pH	Cool 4 deg C	ATGG01	A12	06/10/2025	9040C
Q2292-02	SW-2	Water	pH	Cool 4 deg C	ATGG01	A12	06/10/2025	9040C
Q2293-01	SW-3	Water	pH	Cool 4 deg C	ATGG01	A12	06/10/2025	9040C

Date/Time 06/12/25 08:10
 Raw Sample Received by: SP WC
 Raw Sample Relinquished by: CP SM

Date/Time 06-12-25 13:30
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: SP WC