

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: ATG-GREENVILLE AEC **Level:** LOW **SDG No.:** Q2292
Contract: ATGG01 **Lab Code:** CHEM **Case No.:** Q2292 **SAS No.:** Q2292
Matrix: Water **Sample ID:** Q2264-04 **Client ID:** EF-WWDUP
Percent Solids for Sample: NA **Duplicate ID** Q2264-04DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Copper	ug/L	20	1250		1250		0		P
Lead	ug/L	20	21.8		20.8		5		P
Zinc	ug/L	20	237		236		0		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: ATG-GREENVILLE AEC **Level:** LOW **SDG No.:** Q2292
Contract: ATGG01 **Lab Code:** CHEM **Case No.:** Q2292 **SAS No.:** Q2292
Matrix: Water **Sample ID:** Q2264-04MS **Client ID:** EF-WWMSD
Percent Solids for Sample: NA **Duplicate ID** Q2264-04MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Copper	ug/L	20	1410		1380		2		P
Lead	ug/L	20	483		477		1		P
Zinc	ug/L	20	341		338		1		P