

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: ATG-GREENVILLE AEC **SDG No.:** Q2292
Contract: ATGG01 **Lab Code:** CHEM **Case No.:** Q2292 **SAS No.:** Q2292

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168444BS							
Copper	ug/L	150	168		112	85 - 115	P
Lead	ug/L	500	471		94	85 - 115	P
Zinc	ug/L	100	104		104	85 - 115	P