



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ATG-GREENVILLE AEC

SDG No.: Q2293

Contract: ATGG01

Lab Code: CHEM

Case No.: Q2293

SAS No.: Q2293

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Copper	4.60	+/-10	U	20.0	P	06/17/2025	12:46	LB136187
	Lead	2.30	+/-6	U	12.0	P	06/17/2025	12:46	LB136187
	Zinc	3.50	+/-20	U	40.0	P	06/17/2025	12:46	LB136187

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ATG-GREENVILLE AEC SDG No.: Q2293
Contract: ATGG01 Lab Code: CHEM Case No.: Q2293 SAS No.: Q2293

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Copper	4.60	+/-10	U	20.0	P	06/17/2025	13:48	LB136187
	Lead	2.30	+/-6	U	12.0	P	06/17/2025	13:48	LB136187
	Zinc	3.50	+/-20	U	40.0	P	06/17/2025	13:48	LB136187
CCB02	Copper	4.60	+/-10	U	20.0	P	06/17/2025	14:50	LB136187
	Lead	2.30	+/-6	U	12.0	P	06/17/2025	14:50	LB136187
	Zinc	3.50	+/-20	U	40.0	P	06/17/2025	14:50	LB136187
CCB03	Copper	4.60	+/-10	U	20.0	P	06/17/2025	15:42	LB136187
	Lead	2.30	+/-6	U	12.0	P	06/17/2025	15:42	LB136187
	Zinc	3.50	+/-20	U	40.0	P	06/17/2025	15:42	LB136187
CCB04	Copper	4.60	+/-10	U	20.0	P	06/17/2025	16:34	LB136187
	Lead	2.30	+/-6	U	12.0	P	06/17/2025	16:34	LB136187
	Zinc	3.50	+/-20	U	40.0	P	06/17/2025	16:34	LB136187
CCB05	Copper	4.60	+/-10	U	20.0	P	06/17/2025	17:41	LB136187
	Lead	2.94	+/-6	J	12.0	P	06/17/2025	17:41	LB136187
	Zinc	3.50	+/-20	U	40.0	P	06/17/2025	17:41	LB136187
CCB06	Copper	4.60	+/-10	U	20.0	P	06/17/2025	18:42	LB136187
	Lead	2.30	+/-6	U	12.0	P	06/17/2025	18:42	LB136187
	Zinc	3.50	+/-20	U	40.0	P	06/17/2025	18:42	LB136187
CCB07	Copper	4.60	+/-10	U	20.0	P	06/17/2025	19:30	LB136187
	Lead	2.30	+/-6	U	12.0	P	06/17/2025	19:30	LB136187
	Zinc	3.50	+/-20	U	40.0	P	06/17/2025	19:30	LB136187
CCB08	Copper	4.60	+/-10	U	20.0	P	06/17/2025	20:22	LB136187
	Lead	4.53	+/-6	J	12.0	P	06/17/2025	20:22	LB136187
	Zinc	4.70	+/-20	J	40.0	P	06/17/2025	20:22	LB136187

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ATG-GREENVILLE AEC

SDG No.: Q2293

Contract: ATGG01

Lab Code: CHEM

Case No.: Q2293

SAS No.: Q2293

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
-----------	---------	----------------	---------------------	--------------	------	---	------------------	------------------	---------------