

metals
- 5a -
MATRIX SPIKE SUMMARY

client: ATG-GREENVILLE AEC **level:** low **sdg no.:** Q2293
contract: ATGG01 **lab code:** CHEM **case no.:** Q2293 **sas no.:** Q2293
matrix: Water **sample id:** Q2264-04 **client id:** EF-WWMS
Percent Solids for Sample: NA **Spiked ID:** Q2264-04MS **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	75 - 125	1410		1250		150	106		P
Lead	ug/L	75 - 125	483		21.8		500	92		P
Zinc	ug/L	75 - 125	341		237		100	104		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: <u>ATG-GREENVILLE AEC</u>	level: <u>low</u>	sdg no.: <u>Q2293</u>
contract: <u>ATGG01</u>	lab code: <u>CHEM</u>	case no.: <u>Q2293</u> sas no.: <u>Q2293</u>
matrix: <u>Water</u>	sample id: <u>Q2264-04</u>	client id: <u>EF-WWMSD</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q2264-04MSD</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	75 - 125	1380		1250		150	83		P
Lead	ug/L	75 - 125	477		21.8		500	91		P
Zinc	ug/L	75 - 125	338		237		100	102		P