



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

## **CASE NARRATIVE**

### **ATG-GREENVILLE AEC**

**Project Name: AEC-2025-0013- 91 Fulton, White Plains Bus Co**

**Project # N/A**

**Order ID # Q2294**

**Test Name: COD**

#### **A. Number of Samples and Date of Receipt:**

1 Water sample was received on 06/11/2025.

#### **B. Parameters:**

According to the Chain of Custody document, the following analyses were requested: COD. This data package contains results for COD.

#### **C. Analytical Techniques:**

The analysis of COD was based on method SM5220 D.

#### **D. QA/ QC Samples:**

The Holding Times were met for all analysis.

The Blank Spike met requirements for all samples.

The Duplicate analysis met criteria for all samples.

The Matrix Spike analysis met criteria for all samples.

The Matrix Spike Duplicate analysis met criteria for all samples.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

#### **E. Additional Comments:**

---

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature\_\_\_\_\_