

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2299 SAS No.: Q2299 SDG NO.: Q2299

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 06/14/2025

Lab File ID: BN037259.D Time Analyzed: 13:23

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	1044	7.575	2562	10.35	1268	14.22
UPPER LIMIT	2088	8.075	5124	10.851	2536	14.724
LOWER LIMIT	522	7.075	1281	9.851	634	13.724
EPA SAMPLE NO.						
01 PB168458BS	1588	7.58	3911	10.35	1944	14.22
02 DUP02-20250610	856	7.58	2177	10.35	1128	14.22
03 TT174I1-20250610	939	7.58	2285	10.36	1178	14.22
04 RE134D4-20250610	977	7.58	2442	10.35	1265	14.22
05 RE134D3-20250610	1012	7.58	2458	10.35	1295	14.22
06 TT190D1-20250611	1132	7.58	2844	10.35	1545	14.22
07 RW11-MW01I-20250611	1325	7.58	3321	10.35	1702	14.22
08 RW11-MW01S-20250611	1398	7.58	3634	10.35	1888	14.22
09 RE134D1-20250611	1239	7.58	3144	10.35	1632	14.22
10 TT190D2-20250611	1248	7.58	3038	10.35	1624	14.22

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.