

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2299 SAS No.: Q2299 SDG NO.: Q2299

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 06/16/2025

Lab File ID: BN037284.D Time Analyzed: 10:23

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	1056	7.575	2600	10.35	1364	14.22
UPPER LIMIT	2112	8.075	5200	10.851	2728	14.724
LOWER LIMIT	528	7.075	1300	9.851	682	13.724
EPA SAMPLE NO.						
01 RE131D2-20250610DL	771	7.58	1898	10.36	1101	14.22
02 DUP02-20250610DL	691	7.58	1765	10.36	1009	14.22
03 RE134D3-20250610DL	929	7.58	2314	10.36	1314	14.22

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.