

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q2299 SAS No.: Q2299 SDG No.: Q2299

Instrument ID: MSVOA_N Calibration Date/Time: 06/18/2025 10:21

Lab File ID: VN087074.D Init. Calib. Date(s): 06/06/2025 06/06/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 12:44 14:49

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.499	0.516		3.41	20
Chloromethane	0.644	0.551	0.1	-14.44	20
Vinyl Chloride	0.664	0.705		6.18	20
Bromomethane	0.372	0.373		0.27	20
Chloroethane	0.429	0.449		4.66	20
Trichlorofluoromethane	0.868	0.904		4.15	20
1,1,2-Trichlorotrifluoroethane	0.545	0.568		4.22	20
1,1-Dichloroethene	0.557	0.578		3.77	20
Acetone	0.355	0.348		-1.97	20
Carbon Disulfide	1.539	1.665		8.19	20
Methyl tert-butyl Ether	2.015	1.968		-2.33	20
Methyl Acetate	1.035	0.810		-21.74	20
Methylene Chloride	0.665	0.647		-2.71	20
trans-1,2-Dichloroethene	0.619	0.624		0.81	20
1,1-Dichloroethane	1.120	1.103	0.1	-1.52	20
Cyclohexane	1.086	0.982		-9.58	20
2-Butanone	0.577	0.489		-15.25	20
Carbon Tetrachloride	0.433	0.450		3.93	20
cis-1,2-Dichloroethene	0.740	0.736		-0.54	20
Bromochloromethane	0.550	0.488		-11.27	20
Chloroform	1.118	1.081		-3.31	20
1,1,1-Trichloroethane	0.951	0.924		-2.84	20
Methylcyclohexane	0.605	0.545		-9.92	20
Benzene	1.444	1.468		1.66	20
1,2-Dichloroethane	0.438	0.435		-0.69	20
Trichloroethene	0.342	0.365		6.72	20
1,2-Dichloropropane	0.351	0.360		2.56	20
Bromodichloromethane	0.480	0.491		2.29	20
4-Methyl-2-Pentanone	0.543	0.504		-7.18	20
Toluene	0.882	0.913		3.52	20
t-1,3-Dichloropropene	0.537	0.563		4.84	20
cis-1,3-Dichloropropene	0.574	0.605		5.4	20
1,1,2-Trichloroethane	0.340	0.346		1.76	20
2-Hexanone	0.350	0.340		-2.86	20
Dibromochloromethane	0.354	0.375		5.93	20
1,2-Dibromoethane	0.348	0.363		4.31	20
Tetrachloroethene	0.316	0.325		2.85	20
Chlorobenzene	1.103	1.139	0.3	3.26	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q2299 SAS No.: Q2299 SDG No.: Q2299

Instrument ID: MSVOA_N Calibration Date/Time: 06/18/2025 10:21

Lab File ID: VN087074.D Init. Calib. Date(s): 06/06/2025 06/06/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 12:44 14:49

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.900	1.933		1.74	20
m/p-Xylenes	0.727	0.764		5.09	20
o-Xylene	0.696	0.732		5.17	20
Styrene	1.192	1.262		5.87	20
Bromoform	0.263	0.290	0.1	10.27	20
Isopropylbenzene	3.643	3.597		-1.26	20
1,1,2,2-Tetrachloroethane	1.234	1.231	0.3	-0.24	20
1,3-Dichlorobenzene	1.644	1.705		3.71	20
1,4-Dichlorobenzene	1.676	1.723		2.8	20
1,2-Dichlorobenzene	1.579	1.614		2.22	20
1,2-Dibromo-3-Chloropropane	0.295	0.276		-6.44	20
1,2,4-Trichlorobenzene	1.008	0.900		-10.71	20
1,2,3-Trichlorobenzene	1.002	0.863		-13.87	20
1,2-Dichloroethane-d4	0.669	0.621		-7.18	20
Dibromofluoromethane	0.296	0.316		6.76	20
Toluene-d8	1.173	1.162		-0.94	20
4-Bromofluorobenzene	0.436	0.447		2.52	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.