



Lab Name:	<u>CHEMTECH</u>	Contract:	<u>AECO15</u>				
Lab Code:	<u>CHEM</u>	Case No.:	<u>Q2299</u>	SAS No.:	<u>Q2299</u>	SDG NO.:	<u>Q2299</u>
Lab File ID:	<u>VN087017.D</u>	Date Analyzed:	<u>06/16/2025</u>				
Instrument ID:	<u>MSVOA_N</u>	Time Analyzed:	<u>09:48</u>				
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u>	(mm)	Heated Purge:	<u>(Y/N)</u>	<u>N</u>

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	190027	8.23	334516	9.11	287112	11.87
UPPER LIMIT	380054	8.73	669032	9.606	574224	12.365
LOWER LIMIT	95013.5	7.73	167258	8.606	143556	11.365
EPA SAMPLE NO.						
RE117D1-20250609	248893	8.23	491697	9.11	447651	11.87
TT191D2-20250609	210130	8.23	419506	9.11	373250	11.87
RW10-MW01S-20250609	213340	8.23	416537	9.11	374911	11.87
RW10-MW01D-20250609	291897	8.23	581111	9.11	515699	11.87
RW10A-MW01S-20250609	204665	8.24	406666	9.11	371678	11.87
RW10A-MW01I-20250609	211245	8.24	419153	9.11	374756	11.87
TT158I1-20250610	260360	8.23	512497	9.11	459773	11.87
DUP01-20250610	202159	8.24	394437	9.11	358386	11.87
RE131D2-20250610	274702	8.23	538528	9.11	482289	11.87
DUP02-20250610	301197	8.23	600709	9.11	535463	11.87
TT174I1-20250610	203501	8.23	412992	9.11	371450	11.87
RE134D4-20250610	201975	8.23	409116	9.11	370070	11.87
RE134D3-20250610	194213	8.24	382353	9.11	344228	11.87
TT190D1-20250611	278563	8.23	553254	9.11	491886	11.87
RW11-MW01I-20250611	254338	8.23	501551	9.11	446146	11.87
RW11-MW01S-20250611	278071	8.24	562910	9.11	508781	11.87
RE134D1-20250611	205063	8.24	411667	9.11	371283	11.87
TT190D2-20250611	194784	8.23	389919	9.11	348469	11.87
VN0616WBL01	283331	8.23	556820	9.11	493060	11.87
VN0616WBS01	198479	8.23	351201	9.11	304875	11.87

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
 Lab Code: CHEM Case No.: Q2299 SAS No.: Q2299 SDG NO.: Q2299
 Lab File ID: VN087017.D Date Analyzed: 06/16/2025
 Instrument ID: MSVOA_N Time Analyzed: 09:48
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	190027	8.23	334516	9.11	287112	11.87
UPPER LIMIT	380054	8.73	669032	9.606	574224	12.365
LOWER LIMIT	95013.5	7.73	167258	8.606	143556	11.365
EPA SAMPLE NO.						

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.