

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15  
 Lab Code: CHEM Case No.: Q2299 SAS No.: Q2299 SDG NO.: Q2299  
 Lab File ID: VN087074.D Date Analyzed: 06/18/2025  
 Instrument ID: MSVOA\_N Time Analyzed: 10:21  
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

|                    | IS1<br>AREA # | RT #  | IS2<br>AREA # | RT #  | IS3<br>AREA # | RT #   |
|--------------------|---------------|-------|---------------|-------|---------------|--------|
| 12 HOUR STD        | 217200        | 8.23  | 377846        | 9.11  | 330017        | 11.87  |
| UPPER LIMIT        | 434400        | 8.729 | 755692        | 9.606 | 660034        | 12.365 |
| LOWER LIMIT        | 108600        | 7.729 | 188923        | 8.606 | 165009        | 11.365 |
| EPA SAMPLE NO.     |               |       |               |       |               |        |
| RE117D2-20250609   | 190617        | 8.24  | 387436        | 9.11  | 350863        | 11.87  |
| TT191D1-20250609MS | 197403        | 8.23  | 358053        | 9.11  | 319487        | 11.87  |
| VN0618WBL01        | 270298        | 8.23  | 541331        | 9.11  | 486295        | 11.87  |
| VN0618WBS01        | 213664        | 8.23  | 379722        | 9.11  | 333122        | 11.87  |

IS1 = Pentafluorobenzene  
 IS2 = 1,4-Difluorobenzene  
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.