

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>CHEMTECH</u>	Contract:	<u>AECO15</u>				
Lab Code:	<u>CHEM</u>	Case No.:	<u>Q2299</u>	SAS No.:	<u>Q2299</u>	SDG NO.:	<u>Q2299</u>
Lab File ID:	<u>VN087017.D</u>	Date Analyzed:	<u>06/16/2025</u>				
Instrument ID:	<u>MSVOA_N</u>	Time Analyzed:	<u>09:48</u>				
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u>	(mm)	Heated Purge:	<u>(Y/N)</u>	<u>N</u>

	IS4 AREA #	RT #				
12 HOUR STD	143033	13.788				
UPPER LIMIT	286066	14.288				
LOWER LIMIT	71516.5	13.288				
EPA SAMPLE NO.						
RE117D1-20250609	215929	13.79				
TT191D2-20250609	177105	13.79				
RW10-MW01S-20250609	175863	13.79				
RW10-MW01D-20250609	249769	13.79				
RW10A-MW01S-20250609	176246	13.79				
RW10A-MW01I-20250609	175252	13.79				
TT158I1-20250610	216825	13.79				
DUP01-20250610	168181	13.79				
RE131D2-20250610	227199	13.79				
DUP02-20250610	254156	13.79				
TT174I1-20250610	175583	13.79				
RE134D4-20250610	169597	13.79				
RE134D3-20250610	162344	13.79				
TT190D1-20250611	231060	13.79				
RW11-MW01I-20250611	210117	13.79				
RW11-MW01S-20250611	240142	13.79				
RE134D1-20250611	176304	13.79				
TT190D2-20250611	163211	13.79				
VN0616WBL01	235789	13.79				
VN0616WBS01	154394	13.79				

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
 Lab Code: CHEM Case No.: Q2299 SAS No.: Q2299 SDG NO.: Q2299
 Lab File ID: VN087017.D Date Analyzed: 06/16/2025
 Instrument ID: MSVOA_N Time Analyzed: 09:48
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	143033	13.788				
UPPER LIMIT	286066	14.288				
LOWER LIMIT	71516.5	13.288				
EPA SAMPLE NO.						

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.