

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
 Lab Code: CHEM Case No.: Q2299 SAS No.: Q2299 SDG NO.: Q2299
 Lab File ID: VN087045.D Date Analyzed: 06/17/2025
 Instrument ID: MSVOA_N Time Analyzed: 09:25
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	137590	13.788				
UPPER LIMIT	275180	14.288				
LOWER LIMIT	68795	13.288				
EPA SAMPLE NO.						
TT191D1-20250609MSD	123202	13.79				
TT191D1-20250609	171664	13.79				
RE134D4-20250610DL	219879	13.79				
RE134D3-20250610DL	255285	13.79				
TT190D1-20250611DL	258644	13.79				
VN0617WBL01	170553	13.79				
VN0617WBS01	122529	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.