

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q2300

Project: Monthly 2025

Run No.: LB136135

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136135BS							
BOD5	mg/L	198	213		107	1	84.6-115.4	06/12/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q2300

Project: Monthly 2025

Run No.: LB136144

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136144BS							
TPH	mg/L	20.0	17.0		85	1	78-114	06/13/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q2300

Project: Monthly 2025

Run No.: LB136144

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136144BSD							
TPH	mg/L	20.0	17.0		85	1	78-114	06/13/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q2300

Project: Monthly 2025

Run No.: LB136145

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136145BS							
TSS	mg/L	550	532		97	1	90-110	06/13/2025