

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB136137

Slope : 99.2

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER, COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	06/12/2025	15:00
2	CAL2	1	Water	NA	NA	20.2	7.01	06/12/2025	15:01
3	CAL3	1	Water	NA	NA	20.2	10.02	06/12/2025	15:05
4	ICV	1	Water	NA	NA	20.2	7.00	06/12/2025	15:09
5	CCV1	1	Water	NA	NA	20.2	2.01	06/12/2025	15:11
6	Q2298-01	1	Solid	20.02	20	21.9	8.31	06/12/2025	17:05
7	Q2301-02	1	Solid	20.03	20	22.3	8.13	06/12/2025	17:10
8	Q2301-02DUP	1	Solid	20.04	20	22.4	8.14	06/12/2025	17:11
9	CCV2	1	Water	NA	NA	20.3	12.02	06/12/2025	17:26

WORKLIST(Hardcopy Internal Chain)

136137

WorkList Name : ph s q2297

WorkList ID : 190149

Department : Wet-Chemistry

Date : 06-12-2025 12:40:05

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2298-01	AU-05-061125	Solid	pH	Cool 4 deg C	PSEG05	D41	06/11/2025	9045D
Q2301-02	WC-URBAN-FILL-C	Solid	pH	Cool 4 deg C	ENTA05	D41	06/11/2025	9045D

Date/Time 06-12-25 13:00

Raw Sample Received by: 88 WC

Raw Sample Relinquished by: [Signature]

Date/Time 16-12-25

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]