

TOTAL SOLIDS - SM2540B

SUPERVISOR: Iwona

ANALYST: jignesh

Date: 06/13/2025

Run Number: LB136155

BalanceID: WC SC-6

OvenID: WC OVEN-1

ThermometerID: WET OVEN#1

TEMP1 IN: 103 °C 06/13/2025 11:00 **TEMP1 OUT:** 104 °C 06/13/2025 12:00
TEMP2 IN: 103 °C 06/13/2025 12:30 **TEMP2 OUT:** 104 °C 06/13/2025 13:30
TEMP3 IN: 103 °C 06/13/2025 16:00 **TEMP3 OUT:** 104 °C 06/16/2025 07:00
TEMP4 IN: 103 °C 06/16/2025 07:30 **TEMP4 OUT:** 104 °C 06/16/2025 09:00

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Dish + Sample Weight (g)	Original weight 1st Dish+Sample weight after Drying @103-@105°C (g)	Constant weight 2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Constant weight Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result %
1	LB136155BL	LB136155BL	80.3671	80.3671	80.3671	80.3671	80.3671	80.3671	0.0000	0
2	Q2301-02	WC-URBAN-FILL-C	78.4184	78.4184	104.6063	99.3138	99.3138	99.3138	20.8954	79.8
3	Q2301-02DUP	WC-URBAN-FILL-CDUP	90.3280	90.3280	130.931	123.3999	123.3999	123.4000	33.0719	81.5

A = Final Empty Dish Weight (g)

B = Dish + Sample Weight (g)

C = Final Dish+Sample weight after Drying @103-@105°C (g)

$$\text{Result \%} = (C - A) * 100 / (B - A)$$

WORKLIST(Hardcopy Internal Chain)

VB 136149 VAF6 VB 136155

WorkList Name : ts q2301 solids

WorkList ID : 190193

Department : Wet-Chemistry

Date : 06-13-2025 12:31:15

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2301-02	WC-URBAN-FILL-C	Solid	TS	Cool 4 deg C	ENTA05	D41	06/11/2025	SM2540 B

Date/Time 06/13/25 12:40

Raw Sample Received by: SP

Raw Sample Relinquished by: SP

Date/Time 06/13/25

Raw Sample Received by: SP

Raw Sample Relinquished by: SP