

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q2301

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB136136

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Corrosivity	ICV	pH	7.00	7	100	90-110	06/12/2025
Sample ID: Corrosivity	CCV1	pH	2.01	2.00	101	90-110	06/12/2025
Sample ID: Corrosivity	CCV2	pH	12.02	12.00	100	90-110	06/12/2025
Sample ID: Corrosivity	CCV3	pH	2.01	2.00	101	90-110	06/12/2025

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q2301

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB136137

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
pH		pH	7.00	7	100	90-110	06/12/2025
Sample ID:	CCV1						
pH		pH	2.01	2.00	101	90-110	06/12/2025
Sample ID:	CCV2						
pH		pH	12.02	12.00	100	90-110	06/12/2025

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q2301

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB136162

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Reactive Cyanide	mg/L	0.094	0.099	95	85-115	06/16/2025
Sample ID: CCV1 Reactive Cyanide	mg/L	0.23	0.25	92	90-110	06/16/2025
Sample ID: CCV2 Reactive Cyanide	mg/L	0.23	0.25	92	90-110	06/16/2025

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q2301

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB136163

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 ASTM Ammonia	mg/L	1	1	100	90-110	06/16/2025
Sample ID: CCV1 ASTM Ammonia	mg/L	1	1	100	90-110	06/16/2025
Sample ID: CCV2 ASTM Ammonia	mg/L	0.97	1	97	90-110	06/16/2025

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q2301

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB136170

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV ASTM COD	mg/L	50.962	50	102	95-105	05/28/2025
Sample ID: CCV1 ASTM COD	mg/L	50.962	50	102	95-105	06/16/2025
Sample ID: CCV2 ASTM COD	mg/L	50.962	50	102	95-105	06/16/2025

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q2301

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB136170

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
---------	-------	--------	------------	---------------	---------------------------	------------------