



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Blank Summary

Client: ENTACT

SDG No.: Q2301

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB136162

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Reactive Cyanide	mg/L	0.0013	0.0025	J	0.00096	0.005	06/16/2025
Sample ID: CCB1 Reactive Cyanide	mg/L	0.0013	0.0025	J	0.00096	0.005	06/16/2025
Sample ID: CCB2 Reactive Cyanide	mg/L	0.0013	0.0025	J	0.00096	0.005	06/16/2025

Initial and Continuing Calibration Blank Summary

Client: ENTACT

SDG No.: Q2301

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB136163

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 ASTM Ammonia	mg/L	< 0.0500	0.0500	U	0.030	0.1	06/16/2025
Sample ID: CCB1 ASTM Ammonia	mg/L	< 0.0500	0.0500	U	0.030	0.1	06/16/2025
Sample ID: CCB2 ASTM Ammonia	mg/L	< 0.0500	0.0500	U	0.030	0.1	06/16/2025

Initial and Continuing Calibration Blank Summary

Client: ENTACT

SDG No.: Q2301

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB136170

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB ASTM COD	mg/L	< 5.0000	5.0000	U	1.50	10	05/28/2025
Sample ID: CCB1 ASTM COD	mg/L	< 5.0000	5.0000	U	1.50	10	06/16/2025
Sample ID: CCB2 ASTM COD	mg/L	< 5.0000	5.0000	U	1.50	10	06/16/2025

Initial and Continuing Calibration Blank Summary

Client: ENTACT

SDG No.: Q2301

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB136170

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
---------	-------	--------	----------------------	--------------	-----	-----	------------------