

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: Q2301

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB136142

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136142BS							
Oil and Grease	mg/Kg	100	94.9		95	1	80-120	06/13/2025

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: Q2301

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB136151

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136151BS							
ASTM Oil and Grease	mg/L	20.0	16.7		84	1	78-114	06/13/2025

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: Q2301

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB136170

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136170BS							
ASTM COD	mg/L	50	48.9		98	1	90-110	06/16/2025

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: Q2301

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB136163

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB168481BS							
ASTM Ammonia	mg/L	1	1.00		100	1	90-110	06/16/2025