


LB136157

7470A

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%RSD
0.0	0.000	0.045	0.045	-145	0.000	-145	0				1
0.2	0.200	0.240	0.040	833	0.0 %	833					20
2.5	2.500	2.460	-0.040	12022	0.0 %	12022					2
5.0	5.000	4.862	-0.138	24128	0.0 %	24128					3
7.5	7.500	7.571	0.071	37777	0.0 %	37777					1
10.0	10.000	10.022	0.022	50129	0.0 %	50129					0

LB136157 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 16 Jun 2025 08:49:39

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-145	-	0.0000	7470A	PPB	16 Jun 2025 09:01:55	S	Std
0.2 - 1	50.2	833	-	0.2000	7470A	PPB	16 Jun 2025 09:04:13	S	Std
2.5 - 1	52.5	12022	-	2.5000	7470A	PPB	16 Jun 2025 09:06:31	S	Std
5.0 - 1	55	24128	-	5.0000	7470A	PPB	16 Jun 2025 09:08:48	S	Std
7.5 - 1	57.5	37777	-	7.5000	7470A	PPB	16 Jun 2025 09:11:06	S	Std
10.0 - 1	516	50129	-	10.0000	7470A	PPB	16 Jun 2025 09:13:25	S	Std
ICV38 - 1	ICV38	17917	3.6297	-	7470A	PPB	16 Jun 2025 09:16:48	U	SMPL
ICB38 - 1	ICB38	-150	0.0445	-	7470A	PPB	16 Jun 2025 09:19:03	U	SMPL
CCV42 - 1	CCV42	24528	4.9416	-	7470A	PPB	16 Jun 2025 09:21:20	U	SMPL
CCB42 - 1	CCB42	-382	-0.0016	-	7470A	PPB	16 Jun 2025 09:23:35	U	SMPL
CRA - 1	CRA	744	0.2219	-	7470A	PPB	16 Jun 2025 09:25:52	U	SMPL
HighStd - 1	HighStd	49716	9.9400	-	7470A	PPB	16 Jun 2025 09:28:08	U	SMPL
ChkStd - 1	ChkStd	33071	6.6370	-	7470A	PPB	16 Jun 2025 09:30:22	U	SMPL
PB168479BL - 1	PBW	-453	-0.0157	-	7470A	PPB	16 Jun 2025 09:32:41	U	SMPL
PB168479BS - 1	LCSW	17905	3.6274	-	7470A	PPB	16 Jun 2025 09:40:00	U	SMPL
Q2287-02 - 1	CONCRETE-SLAB	-217	0.0312	-	7470A	PPB	16 Jun 2025 09:42:17	U	SMPL
Q2295-04 - 1	TP01-MH1-WC	-111	0.0522	-	7470A	PPB	16 Jun 2025 09:44:36	U	SMPL
Q2295-08 - 1	TP02-MH5-WC	-121	0.0502	-	7470A	PPB	16 Jun 2025 09:46:53	U	SMPL
Q2296-04 - 1	WC-1	-97	0.0550	-	7470A	PPB	16 Jun 2025 09:49:10	U	SMPL
Q2296-08 - 1	WC-2	-77	0.0589	-	7470A	PPB	16 Jun 2025 09:51:25	U	SMPL
CCV43 - 1	CCV43	23956	4.8281	-	7470A	PPB	16 Jun 2025 09:53:41	U	SMPL
CCB43 - 1	CCB43	-138	0.0468	-	7470A	PPB	16 Jun 2025 09:55:56	U	SMPL
Q2296-12 - 1	WC-3	65	0.0871	-	7470A	PPB	16 Jun 2025 09:58:14	U	SMPL
Q2296-16 - 1	WC-4	42	0.0826	-	7470A	PPB	16 Jun 2025 10:00:29	U	SMPL
Q2296-20 - 1	WC-5	-99	0.0546	-	7470A	PPB	16 Jun 2025 10:02:45	U	SMPL
Q2296-24 - 1	WC-6	-250	0.0246	-	7470A	PPB	16 Jun 2025 10:05:01	U	SMPL
Q2297-04 - 1	TP-3	-77	0.0589	-	7470A	PPB	16 Jun 2025 10:07:19	U	SMPL
Q2301-03 - 1	WC-URBAN-FILL-C	-77	0.0589	-	7470A	PPB	16 Jun 2025 10:09:36	U	SMPL
Q2310-04 - 1	TP-7	-53	0.0637	-	7470A	PPB	16 Jun 2025 10:11:53	U	SMPL
Q2310-04DUP - 1	TP-7DUP	106	0.0953	-	7470A	PPB	16 Jun 2025 10:14:10	U	SMPL
Q2310-04MS - 1	TP-7MS	17397	3.5265	-	7470A	PPB	16 Jun 2025 10:21:12	U	SMPL
Q2310-04MSD - 1	TP-7MSD	17391	3.5254	-	7470A	PPB	16 Jun 2025 10:23:26	U	SMPL
CCV44 - 1	CCV44	23336	4.7051	-	7470A	PPB	16 Jun 2025 10:25:44	U	SMPL
CCB44 - 1	CCB44	-373	0.0002	-	7470A	PPB	16 Jun 2025 10:28:02	U	SMPL
PB168428TB - 1	PB168428TB	-82	0.0580	-	7470A	PPB	16 Jun 2025 10:30:20	U	SMPL
Q2310-04LX5 - 1		-20	0.0703	-	7470A	PPB	16 Jun 2025 10:32:36	U	SMPL
Q2310-04A - 1		17172	3.4819	-	7470A	PPB	16 Jun 2025 10:34:52	U	SMPL
CCV45 - 1	CCV45	23565	4.7505	-	7470A	PPB	16 Jun 2025 10:40:01	U	SMPL
CCB45 - 1	CCB45	-281	0.0185	-	7470A	PPB	16 Jun 2025 10:42:16	U	SMPL