



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: ENTACT **SDG No.:** Q2301
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q2301 **SAS No.:** Q2301
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.22	0.2	111	70 - 130	CV	06/16/2025	09:25	LB136157
CRI01	Arsenic	16.3	20.0	82	65 - 135	P	06/16/2025	14:53	lb136164
	Barium	101	100	101	65 - 135	P	06/16/2025	14:53	lb136164
	Cadmium	5.66	6.0	94	65 - 135	P	06/16/2025	14:53	lb136164
	Chromium	9.79	10.0	98	65 - 135	P	06/16/2025	14:53	lb136164
	Copper	23.0	20.0	115	65 - 135	P	06/16/2025	14:53	lb136164
	Lead	11.5	12.0	96	65 - 135	P	06/16/2025	14:53	lb136164
	Nickel	38.8	40.0	97	65 - 135	P	06/16/2025	14:53	lb136164
	Selenium	17.6	20.0	88	65 - 135	P	06/16/2025	14:53	lb136164
	Silver	11.0	10.0	110	65 - 135	P	06/16/2025	14:53	lb136164
	Zinc	41.7	40.0	104	65 - 135	P	06/16/2025	14:53	lb136164