



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	ENTACT				SDG No.:	Q2301				
Contract:	ENTA05	Lab Code:	CHEM		Case No.:	Q2301	SAS No.:	Q2301		
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number	
ICB38	Mercury	0.076	+/-0.2	U	0.20	CV	06/16/2025	09:19	LB136157	

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ENTACT

SDG No.: Q2301

Contract: ENTA05

Lab Code: CHEM

Case No.: Q2301

SAS No.: Q2301

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB42	Mercury	0.076	+/-0.2	U	0.20	CV	06/16/2025	09:23	LB136157
CCB43	Mercury	0.076	+/-0.2	U	0.20	CV	06/16/2025	09:55	LB136157
CCB44	Mercury	0.076	+/-0.2	U	0.20	CV	06/16/2025	10:28	LB136157
CCB45	Mercury	0.076	+/-0.2	U	0.20	CV	06/16/2025	10:42	LB136157

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ENTACT **SDG No.:** Q2301
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q2301 **SAS No.:** Q2301

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	5.12	+/-10	U	20.0	P	06/16/2025	14:49	lb136164
	Barium	14.6	+/-50	U	100	P	06/16/2025	14:49	lb136164
	Cadmium	0.50	+/-3	U	6.00	P	06/16/2025	14:49	lb136164
	Chromium	2.12	+/-5	U	10.0	P	06/16/2025	14:49	lb136164
	Copper	4.60	+/-10	U	20.0	P	06/16/2025	14:49	lb136164
	Lead	2.30	+/-6	U	12.0	P	06/16/2025	14:49	lb136164
	Nickel	3.06	+/-20	U	40.0	P	06/16/2025	14:49	lb136164
	Selenium	9.64	+/-10	U	20.0	P	06/16/2025	14:49	lb136164
	Silver	1.62	+/-5	U	10.0	P	06/16/2025	14:49	lb136164
	Zinc	3.50	+/-20	U	40.0	P	06/16/2025	14:49	lb136164

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	ENTACT		SDG No.:	Q2301	
Contract:	ENTA05	Lab Code:	CHEM	Case No.:	Q2301
				SAS No.:	Q2301

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	5.12	+/-10	U	20.0	P	06/16/2025	15:30	lb136164
	Barium	14.6	+/-50	U	100	P	06/16/2025	15:30	lb136164
	Cadmium	1.35	+/-3	J	6.00	P	06/16/2025	15:30	lb136164
	Chromium	2.12	+/-5	U	10.0	P	06/16/2025	15:30	lb136164
	Copper	4.60	+/-10	U	20.0	P	06/16/2025	15:30	lb136164
	Lead	3.27	+/-6	J	12.0	P	06/16/2025	15:30	lb136164
	Nickel	3.06	+/-20	U	40.0	P	06/16/2025	15:30	lb136164
	Selenium	9.64	+/-10	U	20.0	P	06/16/2025	15:30	lb136164
	Silver	1.62	+/-5	U	10.0	P	06/16/2025	15:30	lb136164
	Zinc	5.12	+/-20	J	40.0	P	06/16/2025	15:30	lb136164
CCB02	Arsenic	5.12	+/-10	U	20.0	P	06/16/2025	16:34	lb136164
	Barium	14.6	+/-50	U	100	P	06/16/2025	16:34	lb136164
	Cadmium	0.50	+/-3	U	6.00	P	06/16/2025	16:34	lb136164
	Chromium	2.12	+/-5	U	10.0	P	06/16/2025	16:34	lb136164
	Copper	4.60	+/-10	U	20.0	P	06/16/2025	16:34	lb136164
	Lead	2.30	+/-6	U	12.0	P	06/16/2025	16:34	lb136164
	Nickel	3.06	+/-20	U	40.0	P	06/16/2025	16:34	lb136164
	Selenium	9.64	+/-10	U	20.0	P	06/16/2025	16:34	lb136164
	Silver	1.62	+/-5	U	10.0	P	06/16/2025	16:34	lb136164
	Zinc	3.50	+/-20	U	40.0	P	06/16/2025	16:34	lb136164
CCB03	Arsenic	5.12	+/-10	U	20.0	P	06/16/2025	17:49	lb136164
	Barium	14.6	+/-50	U	100	P	06/16/2025	17:49	lb136164
	Cadmium	0.50	+/-3	U	6.00	P	06/16/2025	17:49	lb136164
	Chromium	2.12	+/-5	U	10.0	P	06/16/2025	17:49	lb136164
	Copper	4.60	+/-10	U	20.0	P	06/16/2025	17:49	lb136164
	Lead	2.30	+/-6	U	12.0	P	06/16/2025	17:49	lb136164
	Nickel	3.06	+/-20	U	40.0	P	06/16/2025	17:49	lb136164
	Selenium	9.64	+/-10	U	20.0	P	06/16/2025	17:49	lb136164
	Silver	1.62	+/-5	U	10.0	P	06/16/2025	17:49	lb136164
	Zinc	3.50	+/-20	U	40.0	P	06/16/2025	17:49	lb136164
CCB04	Arsenic	5.12	+/-10	U	20.0	P	06/16/2025	18:33	lb136164
	Barium	14.6	+/-50	U	100	P	06/16/2025	18:33	lb136164
	Cadmium	0.50	+/-3	U	6.00	P	06/16/2025	18:33	lb136164
	Chromium	2.12	+/-5	U	10.0	P	06/16/2025	18:33	lb136164
	Copper	4.60	+/-10	U	20.0	P	06/16/2025	18:33	lb136164
	Lead	2.30	+/-6	U	12.0	P	06/16/2025	18:33	lb136164
	Nickel	3.06	+/-20	U	40.0	P	06/16/2025	18:33	lb136164
	Selenium	9.64	+/-10	U	20.0	P	06/16/2025	18:33	lb136164
	Silver	1.62	+/-5	U	10.0	P	06/16/2025	18:33	lb136164

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>ENTACT</u>		SDG No.: <u>Q2301</u>							
Contract: <u>ENTA05</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q2301</u>	SAS No.: <u>Q2301</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Zinc	3.50	+/-20	U	40.0	P	06/16/2025	18:33	lb136164
CCB05	Arsenic	5.12	+/-10	U	20.0	P	06/16/2025	19:21	lb136164
	Barium	14.6	+/-50	U	100	P	06/16/2025	19:21	lb136164
	Cadmium	3.56	+/-3	J*	6.00	P	06/16/2025	19:21	lb136164
	Chromium	2.12	+/-5	U	10.0	P	06/16/2025	19:21	lb136164
	Copper	4.60	+/-10	U	20.0	P	06/16/2025	19:21	lb136164
	Lead	5.87	+/-6	J	12.0	P	06/16/2025	19:21	lb136164
	Nickel	3.47	+/-20	J	40.0	P	06/16/2025	19:21	lb136164
	Selenium	9.64	+/-10	U	20.0	P	06/16/2025	19:21	lb136164
	Silver	1.62	+/-5	U	10.0	P	06/16/2025	19:21	lb136164
	Zinc	3.50	+/-20	U	40.0	P	06/16/2025	19:21	lb136164
CCB06	Arsenic	5.12	+/-10	U	20.0	P	06/16/2025	20:08	lb136164
	Barium	14.6	+/-50	U	100	P	06/16/2025	20:08	lb136164
	Cadmium	0.50	+/-3	U	6.00	P	06/16/2025	20:08	lb136164
	Chromium	2.12	+/-5	U	10.0	P	06/16/2025	20:08	lb136164
	Copper	4.60	+/-10	U	20.0	P	06/16/2025	20:08	lb136164
	Lead	2.30	+/-6	U	12.0	P	06/16/2025	20:08	lb136164
	Nickel	3.06	+/-20	U	40.0	P	06/16/2025	20:08	lb136164
	Selenium	9.64	+/-10	U	20.0	P	06/16/2025	20:08	lb136164
	Silver	1.62	+/-5	U	10.0	P	06/16/2025	20:08	lb136164
	Zinc	3.50	+/-20	U	40.0	P	06/16/2025	20:08	lb136164
CCB07	Arsenic	5.12	+/-10	U	20.0	P	06/16/2025	20:56	lb136164
	Barium	14.6	+/-50	U	100	P	06/16/2025	20:56	lb136164
	Cadmium	0.50	+/-3	U	6.00	P	06/16/2025	20:56	lb136164
	Chromium	2.12	+/-5	U	10.0	P	06/16/2025	20:56	lb136164
	Copper	4.60	+/-10	U	20.0	P	06/16/2025	20:56	lb136164
	Lead	2.30	+/-6	U	12.0	P	06/16/2025	20:56	lb136164
	Nickel	3.06	+/-20	U	40.0	P	06/16/2025	20:56	lb136164
	Selenium	9.64	+/-10	U	20.0	P	06/16/2025	20:56	lb136164
	Silver	1.62	+/-5	U	10.0	P	06/16/2025	20:56	lb136164
	Zinc	3.50	+/-20	U	40.0	P	06/16/2025	20:56	lb136164
CCB08	Arsenic	5.12	+/-10	U	20.0	P	06/16/2025	21:18	lb136164
	Barium	14.6	+/-50	U	100	P	06/16/2025	21:18	lb136164
	Cadmium	0.50	+/-3	U	6.00	P	06/16/2025	21:18	lb136164
	Chromium	2.12	+/-5	U	10.0	P	06/16/2025	21:18	lb136164
	Copper	4.60	+/-10	U	20.0	P	06/16/2025	21:18	lb136164
	Lead	2.30	+/-6	U	12.0	P	06/16/2025	21:18	lb136164
	Nickel	3.06	+/-20	U	40.0	P	06/16/2025	21:18	lb136164
	Selenium	9.64	+/-10	U	20.0	P	06/16/2025	21:18	lb136164

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	ENTACT				SDG No.:	Q2301				
Contract:	ENTA05		Lab Code:	CHEM		Case No.:	Q2301		SAS No.:	Q2301
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number	
CCB08	Silver	1.62	+/-5	U	10.0	P	06/16/2025	21:18	lb136164	
	Zinc	3.50	+/-20	U	40.0	P	06/16/2025	21:18	lb136164	

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ENTACT

SDG No.: Q2301

Contract: ENTA05

Lab Code: CHEM

Case No.: Q2301

SAS No.: Q2301

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
-----------	---------	----------------	---------------------	--------------	------	---	------------------	------------------	---------------