

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: ENTACT **SDG No.:** Q2301
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q2301 **SAS No.:** Q2301
ICS Source: EPA **Instrument ID:** P5

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	-2.78			-20	20	06/16/2025	14:59	lb136164
	Barium	7.67	6.0	128	-94	106	06/16/2025	14:59	lb136164
	Cadmium	-1.53	1.0	153	-5	7	06/16/2025	14:59	lb136164
	Chromium	51.2	52.0	98	42	62	06/16/2025	14:59	lb136164
	Copper	-9.05	2.0	452	-18	22	06/16/2025	14:59	lb136164
	Lead	4.50			-12	12	06/16/2025	14:59	lb136164
	Nickel	1.27	2.0	64	-38	42	06/16/2025	14:59	lb136164
	Selenium	-0.94			-20	20	06/16/2025	14:59	lb136164
	Silver	5.28			-10	10	06/16/2025	14:59	lb136164
	Zinc	11.4			-40	40	06/16/2025	14:59	lb136164
ICSAB01	Arsenic	94.4	100	94	85	120	06/16/2025	15:06	lb136164
	Barium	539	540	100	440	640	06/16/2025	15:06	lb136164
	Cadmium	1030	970	106	820	1100	06/16/2025	15:06	lb136164
	Chromium	563	540	104	460	620	06/16/2025	15:06	lb136164
	Copper	495	510	97	430	590	06/16/2025	15:06	lb136164
	Lead	52.7	49.0	108	37	61	06/16/2025	15:06	lb136164
	Nickel	1010	950	106	810	1100	06/16/2025	15:06	lb136164
	Selenium	55.7	46.0	121	26	66	06/16/2025	15:06	lb136164
	Silver	173	200	86	170	230	06/16/2025	15:06	lb136164
	Zinc	1070	950	113	810	1100	06/16/2025	15:06	lb136164