

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: ENTACT **Level:** LOW **SDG No.:** Q2301
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q2301 **SAS No.:** Q2301
Matrix: Water **Sample ID:** Q2310-04 **Client ID:** TP-7DUP
Percent Solids for Sample: NA **Duplicate ID** Q2310-04DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	100	U	100	U			P
Barium	ug/L	20	2270		2100		8		P
Cadmium	ug/L	20	4.56	J	3.99	J	13		P
Chromium	ug/L	20	15.8	J	14.1	J	11		P
Copper	ug/L	20	131		131		0		P
Lead	ug/L	20	34.8	J	34.3	J	1		P
Mercury	ug/L	20	2.00	U	0.95	J	200.0		CV
Nickel	ug/L	20	44.0	J	42.7	J	3		P
Selenium	ug/L	20	100	U	100	U			P
Silver	ug/L	20	50.0	U	50.0	U			P
Zinc	ug/L	20	322		324		1		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: ENTACT **Level:** LOW **SDG No.:** Q2301
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q2301 **SAS No.:** Q2301
Matrix: Water **Sample ID:** Q2310-04MS **Client ID:** TP-7MSD
Percent Solids for Sample: NA **Duplicate ID** Q2310-04MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	4150		4050		2		P
Barium	ug/L	20	3400		3390		0		P
Cadmium	ug/L	20	1060		1030		3		P
Chromium	ug/L	20	2080		2020		3		P
Copper	ug/L	20	1620		1580		3		P
Lead	ug/L	20	4940		4850		2		P
Mercury	ug/L	20	35.3		35.3		0		CV
Nickel	ug/L	20	2620		2570		2		P
Selenium	ug/L	20	9860		9680		2		P
Silver	ug/L	20	364		349		4		P
Zinc	ug/L	20	1320		1310		1		P