

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: ENTACT **SDG No.:** Q2301
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q2301 **SAS No.:** Q2301

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168477BS							
Arsenic	ug/L	4000	4020		100	80 - 120	P
Barium	ug/L	1000	1090		109	80 - 120	P
Cadmium	ug/L	1000	1050		105	80 - 120	P
Chromium	ug/L	2000	2180		109	80 - 120	P
Copper	ug/L	1500	1780		119	80 - 120	P
Lead	ug/L	5000	5190		104	80 - 120	P
Nickel	ug/L	2500	2690		108	80 - 120	P
Selenium	ug/L	10000	10900		109	80 - 120	P
Silver	ug/L	380	409		108	80 - 120	P
Zinc	ug/L	1000	1110		111	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: ENTACT

SDG No.: Q2301

Contract: ENTA05

Lab Code: CHEM

Case No.: Q2301

SAS No.: Q2301

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168479BS Mercury	ug/L	4.0	3.63		91	80 - 120	CV