

Metals
-9 -
ICP SERIAL DILUTIONS

SAMPLE NO.

TP-7L

Lab Name: Chemtech Consulting Group

Contract: ENTA05

Lab Code: CHEM **Lb No.:** lb136164 **Lab Sample ID :** Q2310-04L **SDG No.:** Q2301

Matrix (soil/water): Water **Level (low/med):** LOW

Concentration Units: ug/L

Analyte	Initial Sample Result (I)		Serial Dilution Result (S)		% Difference	Q	M
		C		C			
Arsenic	100	U	500	U			P
Barium	2270		2590		14		P
Cadmium	4.56	J	150	U	100.0		P
Chromium	15.8	J	250	U	100.0		P
Copper	131		162	J	23		P
Lead	34.8	J	300	U	100.0		P
Mercury	2.00	U	10.0	U			CV
Nickel	44.0	J	1000	U	100.0		P
Selenium	100	U	500	U			P
Silver	50.0	U	250	U			P
Zinc	322		316	J	2		P