

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2301 SAS No.: Q2301 SDG NO.: Q2301

Continuing Calib Date: 06/13/2025 Initial Calibration Date(s): 05/19/2025 05/19/2025

Continuing Calib Time: 10:50 Initial Calibration Time(s): 09:42 18:25

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1268-1	(1)	8.70	8.72	8.62	8.82	0.02
Aroclor-1268-2	(2)	8.80	8.81	8.71	8.91	0.01
Aroclor-1268-3	(3)	9.03	9.04	8.94	9.14	0.01
Aroclor-1268-4	(4)	9.44	9.46	9.36	9.56	0.02
Aroclor-1268-5	(5)	9.86	9.88	9.78	9.98	0.02
Tetrachloro-m-xylene		4.50	4.50	4.40	4.60	0.00
Decachlorobiphenyl		10.19	10.22	10.12	10.32	0.03

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2301 SAS No.: Q2301 SDG NO.: Q2301

Continuing Calib Date: 06/13/2025 Initial Calibration Date(s): 05/19/2025 05/19/2025

Continuing Calib Time: 10:50 Initial Calibration Time(s): 09:42 18:25

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1268-1	(1)	7.69	7.70	7.60	7.80	0.02
Aroclor-1268-2	(2)	7.75	7.77	7.67	7.87	0.02
Aroclor-1268-3	(3)	7.95	7.97	7.87	8.07	0.02
Aroclor-1268-4	(4)	8.25	8.27	8.17	8.37	0.02
Aroclor-1268-5	(5)	8.55	8.57	8.47	8.67	0.02
Tetrachloro-m-xylene		3.79	3.80	3.70	3.90	0.01
Decachlorobiphenyl		8.80	8.83	8.73	8.93	0.03