

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2301 SAS No.: Q2301 SDG NO.: Q2301

Continuing Calib Date: 06/13/2025 Initial Calibration Date(s): 05/19/2025 05/19/2025

Continuing Calib Time: 19:31 Initial Calibration Time(s): 09:42 18:25

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.65	5.65	5.55	5.75	0.00
Aroclor-1016-2	(2)	5.67	5.67	5.57	5.77	0.00
Aroclor-1016-3	(3)	5.73	5.74	5.64	5.84	0.01
Aroclor-1016-4	(4)	5.83	5.83	5.73	5.93	0.00
Aroclor-1016-5	(5)	6.12	6.13	6.03	6.23	0.01
Aroclor-1260-1	(1)	7.24	7.25	7.15	7.35	0.01
Aroclor-1260-2	(2)	7.49	7.50	7.40	7.60	0.01
Aroclor-1260-3	(3)	7.85	7.86	7.76	7.96	0.01
Aroclor-1260-4	(4)	8.07	8.09	7.99	8.19	0.02
Aroclor-1260-5	(5)	8.39	8.40	8.30	8.50	0.01
Tetrachloro-m-xylene		4.50	4.50	4.40	4.60	0.00
Decachlorobiphenyl		10.19	10.21	10.11	10.31	0.02

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2301 SAS No.: Q2301 SDG NO.: Q2301

Continuing Calib Date: 06/13/2025 Initial Calibration Date(s): 05/19/2025 05/19/2025

Continuing Calib Time: 19:31 Initial Calibration Time(s): 09:42 18:25

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.87	4.88	4.78	4.98	0.01
Aroclor-1016-2	(2)	4.89	4.90	4.80	5.00	0.01
Aroclor-1016-3	(3)	5.07	5.07	4.97	5.17	0.00
Aroclor-1016-4	(4)	5.11	5.12	5.02	5.22	0.01
Aroclor-1016-5	(5)	5.32	5.33	5.23	5.43	0.01
Aroclor-1260-1	(1)	6.35	6.36	6.26	6.46	0.01
Aroclor-1260-2	(2)	6.54	6.55	6.45	6.65	0.01
Aroclor-1260-3	(3)	6.69	6.71	6.61	6.81	0.02
Aroclor-1260-4	(4)	7.16	7.18	7.08	7.28	0.02
Aroclor-1260-5	(5)	7.41	7.42	7.32	7.52	0.01
Tetrachloro-m-xylene		3.79	3.79	3.69	3.89	0.00
Decachlorobiphenyl		8.80	8.82	8.72	8.92	0.02