

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2301 SAS No.: Q2301 SDG NO.: Q2301

Continuing Calib Date: 06/14/2025 Initial Calibration Date(s): 05/19/2025 05/19/2025

Continuing Calib Time: 09:12 Initial Calibration Time(s): 09:42 18:25

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1254-1	(1)	6.50	6.51	6.41	6.61	0.01
Aroclor-1254-2	(2)	6.71	6.72	6.62	6.82	0.01
Aroclor-1254-3	(3)	7.08	7.09	6.99	7.19	0.01
Aroclor-1254-4	(4)	7.36	7.37	7.27	7.47	0.01
Aroclor-1254-5	(5)	7.77	7.79	7.69	7.89	0.02
Tetrachloro-m-xylene		4.50	4.50	4.40	4.60	0.00
Decachlorobiphenyl		10.19	10.22	10.12	10.32	0.03

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2301 SAS No.: Q2301 SDG NO.: Q2301

Continuing Calib Date: 06/14/2025 Initial Calibration Date(s): 05/19/2025 05/19/2025

Continuing Calib Time: 09:12 Initial Calibration Time(s): 09:42 18:25

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1254-1	(1)	5.67	5.69	5.59	5.79	0.02
Aroclor-1254-2	(2)	5.82	5.83	5.73	5.93	0.01
Aroclor-1254-3	(3)	6.22	6.24	6.14	6.34	0.02
Aroclor-1254-4	(4)	6.45	6.47	6.37	6.57	0.02
Aroclor-1254-5	(5)	6.87	6.88	6.78	6.98	0.02
Tetrachloro-m-xylene		3.79	3.80	3.70	3.90	0.01
Decachlorobiphenyl		8.80	8.83	8.73	8.93	0.03