

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2301 SAS No.: Q2301 SDG NO.: Q2301

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL01 Date Analyzed: 06/13/2025

Lab Sample No.: AR1660CCC500 Data File : PP072898.D Time Analyzed: 09:44

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.648	5.552	5.752	433.430	500.000	-13.3
Aroclor-1016-2	5.669	5.574	5.774	480.470	500.000	-3.9
Aroclor-1016-3	5.731	5.636	5.836	475.320	500.000	-4.9
Aroclor-1016-4	5.829	5.734	5.934	490.530	500.000	-1.9
Aroclor-1016-5	6.121	6.027	6.227	488.970	500.000	-2.2
Aroclor-1260-1	7.239	7.147	7.347	476.790	500.000	-4.6
Aroclor-1260-2	7.492	7.401	7.601	459.490	500.000	-8.1
Aroclor-1260-3	7.850	7.760	7.960	479.490	500.000	-4.1
Aroclor-1260-4	8.074	7.985	8.185	477.040	500.000	-4.6
Aroclor-1260-5	8.393	8.303	8.503	502.580	500.000	0.5
Decachlorobiphenyl	10.191	10.113	10.313	53.180	50.000	6.4
Tetrachloro-m-xylene	4.496	4.396	4.596	50.560	50.000	1.1

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2301 SAS No.: Q2301 SDG NO.: Q2301

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL01 Date Analyzed: 06/13/2025

Lab Sample No.: AR1660CCC500 Data File : PP072898.D Time Analyzed: 09:44

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.870	4.777	4.977	527.350	500.000	5.5
Aroclor-1016-2	4.888	4.795	4.995	537.000	500.000	7.4
Aroclor-1016-3	5.065	4.973	5.173	534.350	500.000	6.9
Aroclor-1016-4	5.106	5.015	5.215	541.100	500.000	8.2
Aroclor-1016-5	5.320	5.229	5.429	560.140	500.000	12.0
Aroclor-1260-1	6.352	6.264	6.464	543.650	500.000	8.7
Aroclor-1260-2	6.540	6.453	6.653	519.640	500.000	3.9
Aroclor-1260-3	6.692	6.606	6.806	541.410	500.000	8.3
Aroclor-1260-4	7.163	7.077	7.277	544.390	500.000	8.9
Aroclor-1260-5	7.405	7.319	7.519	549.520	500.000	9.9
Decachlorobiphenyl	8.802	8.723	8.923	60.430	50.000	20.9
Tetrachloro-m-xylene	3.790	3.691	3.891	53.970	50.000	7.9