

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: ENTA05
Lab Code: CHEM **Case No.:** Q2301 **SAS No.:** Q2301 **SDG NO.:** Q2301
Instrument ID: ECD_S **Calibration Date(s):** 06/16/2025 06/16/2025
Calibration Times: 09:56 11:33

GC Column: RTX-CLP **ID:** 0.32 (mm)

LAB FILE ID:		CF 200 = <u>PS030681.D</u>	CF 500 = <u>PS030682.D</u>				
CF 750 = <u>PS030683.D</u>		CF 1000 = <u>PS030684.D</u>	CF 1500 = <u>PS030685.D</u>				
COMPOUND	CF 200	CF 500	CF 750	CF 1000	CF 1500	CF	% RSD
2,4,5-TP(Silvex)	23310400000	20262400000	19069000000	19162800000	18229500000	20006800000	10
2,4-D	4118600000	3402180000	3170830000	3194800000	3115400000	3400360000	12
2,4-DCAA	4709630000	3856360000	3484750000	3444780000	3308550000	3760820000	15

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: ENTA05
Lab Code: CHEM **Case No.:** Q2301 **SAS No.:** Q2301 **SDG NO.:** Q2301
Instrument ID: ECD_S **Calibration Date(s):** 06/16/2025 06/16/2025
Calibration Times: 09:56 11:33

GC Column: RTX-CLP2 **ID:** 0.32 (mm)

LAB FILE ID:		CF 200 =	<u>PS030681.D</u>	CF 500 =	<u>PS030682.D</u>		
CF 750 =		<u>PS030683.D</u>	CF 1000 =	<u>PS030684.D</u>	CF 1500 =	<u>PS030685.D</u>	
COMPOUND	CF 200	CF 500	CF 750	CF 1000	CF 1500	CF	% RSD
2,4,5-TP(Silvex)	17076700000	14594600000	13656900000	13568800000	12875800000	14354600000	11
2,4-D	2067250000	1684360000	1565840000	1560790000	1503980000	1676450000	14
2,4-DCAA	1334550000	1086460000	1006290000	1005060000	966095000	1079690000	14