

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2301 SAS No.: Q2301 SDG NO.: Q2301

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 06/16/2025 06/16/2025

Client Sample No.: CCAL01 Date Analyzed: 06/16/2025

Lab Sample No.: HSTDCCC750 Data File : PS030698.D Time Analyzed: 17:25

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.378	9.277	9.477	724.580	712.500	1.7
2,4-D	8.488	8.388	8.588	728.580	705.000	3.3
2,4-DCAA	7.350	7.250	7.450	712.270	750.000	-5.0

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2301 SAS No.: Q2301 SDG NO.: Q2301

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/16/2025 06/16/2025

Client Sample No.: CCAL01 Date Analyzed: 06/16/2025

Lab Sample No.: HSTDCCC750 Data File : PS030698.D Time Analyzed: 17:25

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.931	9.832	10.032	668.480	712.500	-6.2
2,4-D	9.025	8.926	9.126	667.720	705.000	-5.3
2,4-DCAA	7.768	7.668	7.868	706.810	750.000	-5.8