

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2301 SAS No.: Q2301 SDG NO.: Q2301

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 06/16/2025 06/16/2025

Client Sample No.: CCAL02 Date Analyzed: 06/16/2025

Lab Sample No.: HSTDCCC750 Data File : PS030706.D Time Analyzed: 21:02

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.377	9.277	9.477	713.140	712.500	0.1
2,4-D	8.488	8.388	8.588	699.550	705.000	-0.8
2,4-DCAA	7.350	7.250	7.450	709.590	750.000	-5.4

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2301 SAS No.: Q2301 SDG NO.: Q2301

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/16/2025 06/16/2025

Client Sample No.: CCAL02 Date Analyzed: 06/16/2025

Lab Sample No.: HSTDCCC750 Data File : PS030706.D Time Analyzed: 21:02

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.932	9.832	10.032	680.650	712.500	-4.5
2,4-D	9.026	8.926	9.126	656.830	705.000	-6.8
2,4-DCAA	7.768	7.668	7.868	705.610	750.000	-5.9