

Report of Analysis

Client:	ENTACT		Date Collected:	06/11/25	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	06/11/25	
Client Sample ID:	WC-URBAN-FILL-C		SDG No.:	Q2301	
Lab Sample ID:	Q2301-03		Matrix:	TCLP	
Analytical Method:	8151A		% Solid:	0	Decanted:
Sample Wt/Vol:	100	Units: mL	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	TCLP Herbicide	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	8151A				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PS030702.D	1	06/16/25 08:55	06/16/25 19:26	PB168495

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
94-75-7	2,4-D	0.0092	U	0.0092	0.020	mg/L
93-72-1	2,4,5-TP (Silvex)	0.0078	U	0.0078	0.020	mg/L
SURROGATES						
19719-28-9	2,4-DCAA	448		70 (61) - 130 (136)	90%	SPK: 500

Comments:

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit