

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q2301</u>	Analytical Method:	<u>8151A</u>
Client:	<u>ENTACT</u>	DataFile :	<u>PS030704.D</u>

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID: Q2301-03MSD (Column 1)	WC-URBAN-FILL-CMSD 2,4-D	50	0	55.4	ug/L	111		4		70 (65)	130 (135)	20 (20)
	2,4,5-TP(Silvex)	50	0	51.7	ug/L	103		2		70 (62)	130 (139)	20 (20)
Client Sample ID: Q2301-03MSD (Column 2)	WC-URBAN-FILL-CMSD 2,4-D	50	0	47.6	ug/L	95		5		70 (65)	130 (135)	20 (20)
	2,4,5-TP(Silvex)	50	0	58.9	ug/L	118		2		70 (62)	130 (139)	20 (20)